

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081522\
 Data File : PD071549.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2022 17:52
 Operator : AR\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 16 04:30:58 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081522.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 16 04:30:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.805	3.563	270.7E6	2684.5E6	98.666	96.043
28)	SA Decachlor...	8.000	9.091	275.7E6	926.9E6	97.378	95.675
Target Compounds							
2)	A alpha-BHC	3.315	4.029	439.0E6	3662.9E6	100.110	96.846
3)	MA gamma-BHC...	3.648	4.364	397.0E6	2916.6E6	99.709	96.250
4)	MA Heptachlor	3.995	4.942	417.0E6	2110.1E6	99.222	95.141
5)	MB Aldrin	4.279	5.285	436.5E6	1882.8E6	99.135	95.223
6)	B beta-BHC	3.947	4.571	152.1E6	1086.0E6	97.628	94.722
7)	B delta-BHC	4.181	4.819	401.9E6	2287.2E6	99.814	96.018
8)	B Heptachlo...	4.786	5.717	372.7E6	1528.4E6	98.587	95.150
9)	A Endosulfan I	5.159	6.104	345.5E6	1406.8E6	98.502	95.565
10)	B gamma-Chl...	5.039	5.971	393.9E6	1561.0E6	99.317	96.091
11)	B alpha-Chl...	5.103	6.049	384.6E6	1511.5E6	98.980	95.957
12)	B 4,4'-DDE	5.295	6.222	357.8E6	1436.4E6	99.774	96.239
13)	MA Dieldrin	5.427	6.381	395.1E6	1561.8E6	99.152	96.234
14)	MA Endrin	5.704	6.615	351.3E6	1336.4E6	99.261	96.269
15)	B Endosulfa...	6.000	6.839	319.8E6	1262.1E6	98.551	96.042
16)	A 4,4'-DDD	5.853	6.746	280.2E6	1149.8E6	99.449	96.433
17)	MA 4,4'-DDT	6.107	7.057	308.3E6	1218.6E6	100.206	97.072
18)	B Endrin al...	6.180	6.971	251.4E6	999.3E6	97.817	95.583
19)	B Endosulfa...	6.404	7.204	298.9E6	1140.7E6	98.166	96.210
20)	A Methoxychlor	6.686	7.534	152.9E6	589.0E6	97.040	95.758
21)	B Endrin ke...	6.914	7.695	346.5E6	1237.9E6	97.794	96.172
22)	Mirex	7.101	8.157	274.0E6	880.2E6	97.495	95.675

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081522\
Data File : PD071549.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2022 17:52
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 04:30:58 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081522.M
Quant Title : GC Extractables
QLast Update : Tue Aug 16 04:30:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

