

Chromatographic Resolution Report

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081618\
Data File : PD048962.D
Acq On : 16 Aug 2018 9:20
Operator : AJ\SJ
Sample : RESC20
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e

Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081618CLP.M
Title : GC Extractables
Last Update : Thu Aug 16 13:32:25 2018
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

RT#1	RT#2	Resolution
3.370	3.800	100.00%
3.800	4.078	100.00%
4.078	4.327	100.00%
4.327	4.367	98.98%
4.367	4.525	100.00%
4.525	4.608	100.00%
4.608	5.048	100.00%
5.048	5.270	100.00%
5.270	5.328	100.00%
5.328	5.380	100.00%
5.380	5.495	100.00%
5.495	5.621	100.00%
5.621	5.876	100.00%
5.876	6.007	100.00%
6.007	6.149	100.00%
6.149	6.244	100.00%
6.244	6.318	100.00%
6.318	6.529	100.00%
6.529	6.794	100.00%
6.794	7.017	100.00%
7.017	8.065	100.00%
8.065	9.575	100.00%
9.575	10.151	100.00%

Signal #2

3.968	4.347	100.00%
4.347	4.628	100.00%
4.628	4.788	100.00%
4.788	5.005	100.00%
5.005	5.141	100.00%
5.141	5.445	100.00%
5.445	5.825	100.00%
5.825	6.058	100.00%
6.058	6.131	100.00%
6.131	6.179	94.29%
6.179	6.285	100.00%
6.285	6.433	100.00%
6.433	6.648	100.00%
6.648	6.767	100.00%
6.767	6.850	100.00%
6.850	6.973	100.00%
6.973	7.066	100.00%
7.066	7.195	100.00%
7.195	7.519	100.00%
7.519	7.660	100.00%
7.660	8.992	100.00%
8.992	9.931	100.00%

