

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081619\
 Data File : PD054312.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Aug 2019 12:48
 Operator : SG\AJ
 Sample : PIBLK36
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 16 18:00:01 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081619CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 16 15:23:53 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.312	3.982	14108551	20108138	20.323	20.597
27)	SA Decachlor...	8.005	9.023	27994982	40155384	40.519	41.273


Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081619\
Data File : PD054312.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Aug 2019 12:48
Operator : SG\AJ
Sample : PIBLK36
Misc :
ALS Vial : 19 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 18:00:01 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081619CLP.M
Quant Title : GC Extractables
QLast Update : Fri Aug 16 15:23:53 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm




Instrument :
ECD_D
ClientSampleId :



#1 Tetrachloro-m-xylene


R.T.: 3.312 min
Delta R.T.: 0.000 min
Response: 14108551
Conc: 20.32 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.982 min
Delta R.T.: 0.000 min
Response: 20108138
Conc: 20.60 ng/ml



#27 Decachlorobiphenyl

R.T.: 8.005 min
Delta R.T.: 0.000 min
Response: 27994982
Conc: 40.52 ng/ml



#27 Decachlorobiphenyl

R.T.: 9.023 min
Delta R.T.: -0.001 min
Response: 40155384
Conc: 41.27 ng/ml