

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081619\
 Data File : PD054304.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Aug 2019 10:57
 Operator : SG\AJ
 Sample : INDA201
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA201

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 16 14:26:26 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081619CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 16 14:23:12 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.312	3.983	7295423	10231356	10.322	10.200
27)	SA Decachlor...	8.006	9.024	14267802	20756020	20.750	21.166
Target Compounds							
2)	A alpha-BHC	3.741	4.369	10047205	13498293	9.937	9.855
3)	MA gamma-BHC...	4.018	4.652	10846830	13614537	10.654	9.984
4)	MA Heptachlor	4.306	5.160	9329262	14108517	10.063	10.079
9)	A Endosulfan I	5.317	6.203	8947842	13159656	10.077	10.166
13)	MA Dieldrin	5.557	6.458	19717679	28392694	19.565	19.722
14)	MA Endrin	5.811	6.675	16640718	21439579	19.866	19.714
16)	A 4,4'-DDD	5.946	6.796	10601138	17126552	19.051	19.283
17)	MA 4,4'-DDT	6.183	7.093	11655680	18143112	18.558	18.579
20)	A Methoxychlor	6.733	7.549	28413773	46162426	100.012	99.578

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081619\
Data File : PD054304.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Aug 2019 10:57
Operator : SG\AJ
Sample : INDA201
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
INDA201

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 14:26:26 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081619CLP.M
Quant Title : GC Extractables
QLast Update : Fri Aug 16 14:23:12 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

