Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081619\
Data File : PD054302.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Aug 2019 10:29
Operator : SG\AJ
Sample : INDA101
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument : ECD_D
ClientSampleId : INDA101

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 14:26:13 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081619CLP.M
Quant Title : GC Extractables
QLast Update : Fri Aug 16 14:23:12 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

		Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
		System Monitoring Compounds						
1)	SA	Tetrachlo...	3.311	3.982	3527578	4904523	4.991	4.890
27)	SA	Decachlor...	8.006	9.025	7334300	10755810	10.666	10.968
		Target Compounds						
2)	A	alpha-BHC	3.741	4.368	4854497	6316297	4.801	4.612
3)	MA	gamma-BHC...	4.018	4.651	5823976	6527903	5.721	4.787
4)	MA	Heptachlor	4.306	5.159	4613249	6948573	4.976	4.964
9)	A	Endosulfan I	5.317	6.201	4448567	6423989	5.010	4.962
13)	MA	Dieldrin	5.557	6.457	9700138	14079285	9.625	9.780
14)	MA	Endrin	5.811	6.673	8303426	10505546	9.913	9.660
16)	A	4,4'-DDD	5.946	6.795	5204007	8132301	9.352	9.156
17)	MA	4,4'-DDT	6.183	7.093	5634762	8772336	8.971	8.983
20)	A	Methoxychlor	6.732	7.549	14264796	22577596	50.210	48.702

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081619\
Data File : PD054302.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Aug 2019 10:29
Operator : SG\AJ
Sample : INDA101
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
INDA101

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 14:26:13 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081619CLP.M
Quant Title : GC Extractables
QLast Update : Fri Aug 16 14:23:12 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm