

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081720\
 Data File : PD058975.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Aug 2020 11:48
 Operator : DD\AJ
 Sample : L3611-14
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 05:14:51 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD080620CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Aug 06 11:29:05 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.208	3.866	15221230	23225954	15.739	16.037
27)	SA Decachlor...	7.842	8.861	28191684	31811586	26.765	23.196
Target Compounds							
10)	B trans-Chl...	5.065	5.934	3633446	23073328	3.142m	13.851 #
11)	B cis-Chlor...	5.122	6.019	5100527	8763299	4.383	5.205
12)	B 4,4'-DDE	5.291	6.176	3802573	4481339	3.305	2.828
13)	MA Dieldrin	5.409	6.321	30402595	21313471	25.028	12.733 #
14)	MA Endrin	5.665	6.533	1134537	729485	1.142m	0.538m#
16)	A 4,4'-DDD	5.800	6.666	2422077	7143864	2.568m	5.450 #
17)	MA 4,4'-DDT	6.028	6.961	9051937	6310013	10.065m	4.782m#
19)	B Endosulfa...	6.306	7.097	1750119	1672354	1.710	1.164 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081720\
Data File : PD058975.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Aug 2020 11:48
Operator : DD\AJ
Sample : L3611-14
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:14:51 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD080620CLP.M
Quant Title : GC Extractables
QLast Update : Thu Aug 06 11:29:05 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

