

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081720\
 Data File : PD058986.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Aug 2020 14:25
 Operator : DD\AJ
 Sample : L3611-09
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 06:11:16 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD080620CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Aug 06 11:29:05 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.208	3.866	14288909	20605096	14.775	14.228
27)	SA Decachlor...	7.842	8.862	27146993	27526733	25.773	20.072
Target Compounds							
8)	B Heptachlo...	4.846	5.708	1354912	8837988	1.199	5.308 #
10)	B trans-Chl...	5.067	5.932	8975592	65803710	7.762	39.503m#
11)	B cis-Chlor...	5.122	6.020	16944862	39953056	14.559	23.731 #
12)	B 4,4'-DDE	5.292	6.175	9118562	14761753	7.925	9.315
13)	MA Dieldrin	5.407	6.321	62463538	22251112	51.421	13.293 #
14)	MA Endrin	5.666	6.533	3551790	9151106	3.575m	6.746m#
16)	A 4,4'-DDD	5.803	6.666	4289802	11398889	4.549m	8.696m#
17)	MA 4,4'-DDT	6.028	6.961	26576117	17245216	29.552m	13.068m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081720\
Data File : PD058986.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Aug 2020 14:25
Operator : DD\AJ
Sample : L3611-09
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 06:11:16 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD080620CLP.M
Quant Title : GC Extractables
QLast Update : Thu Aug 06 11:29:05 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

