

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD081721\  
 Data File : PD065010.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 17 Aug 2021 18:44  
 Operator : AR\AJ  
 Sample : PIBLK66  
 Misc :  
 ALS Vial : 19 Sample Multiplier: 1

Instrument :  
 ECD\_D  
 ClientSampleId :  
 PIBLK065

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Aug 18 03:04:04 2021  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD081721CLP.M  
 Quant Title : GC Extractables  
 QLast Update : Wed Aug 18 02:57:59 2021  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 µl  
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

|                             | Compound        | RT#1  | RT#2  | Resp#1   | Resp#2  | ng/ml  | ng/ml  |
|-----------------------------|-----------------|-------|-------|----------|---------|--------|--------|
| -----                       |                 |       |       |          |         |        |        |
| System Monitoring Compounds |                 |       |       |          |         |        |        |
| 1)                          | SA Tetrachlo... | 3.497 | 4.013 | 21024417 | 166.5E6 | 21.940 | 21.716 |
| 27)                         | SA Decachlor... | 8.242 | 9.084 | 43393021 | 195.5E6 | 43.895 | 42.919 |


Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Data\PD081721\  
Data File : PD065010.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 17 Aug 2021 18:44  
Operator : AR\AJ  
Sample : PIBLK66  
Misc :  
ALS Vial : 19 Sample Multiplier: 1

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Aug 18 03:04:04 2021  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_D\Method\PD081721CLP.M  
Quant Title : GC Extractables  
QLast Update : Wed Aug 18 02:57:59 2021  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1  
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Instrument :  
ECD\_D  
ClientSampled :  
PIBLK065