

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081825\
 Data File : PD089966.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Aug 2025 11:53
 Operator : AR\AJ
 Sample : PSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 19 02:23:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081825.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 19 02:15:07 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.548	2.878	143.7E6	1061.4E6	50.000	50.000
28)	SA Decachlor...	9.067	8.065	207.4E6	1249.0E6	50.000	50.000
Target Compounds							
2)	A alpha-BHC	3.997	3.390	303.8E6	1650.3E6	50.000	50.000
3)	MA gamma-BHC...	4.328	3.726	286.8E6	1519.5E6	50.000	50.000
4)	MA Heptachlor	4.926	4.078	278.4E6	1514.2E6	50.000	50.000
5)	MB Aldrin	5.267	4.364	270.5E6	1477.4E6	50.000	50.000
6)	B beta-BHC	4.514	4.021	109.3E6	642.1E6	50.000	50.000
7)	B delta-BHC	4.762	4.258	262.1E6	1531.6E6	50.000	50.000
8)	B Heptachlo...	5.687	4.867	239.7E6	1335.3E6	50.000	50.000
9)	A Endosulfan I	6.071	5.242	228.2E6	1267.8E6	50.000	50.000
10)	B gamma-Chl...	5.943	5.120	242.9E6	1436.0E6	50.000	50.000
11)	B alpha-Chl...	6.024	5.185	243.6E6	1376.6E6	50.000	50.000
12)	B 4,4'-DDE	6.192	5.369	220.2E6	1394.4E6	50.000	50.000
13)	MA Dieldrin	6.344	5.507	244.6E6	1418.7E6	50.000	50.000
14)	MA Endrin	6.571	5.784	208.2E6	1285.5E6	50.000	50.000
15)	B Endosulfa...	6.783	6.075	207.0E6	1221.8E6	50.000	50.000
16)	A 4,4'-DDD	6.702	5.924	172.9E6	1167.3E6	50.000	50.000
17)	MA 4,4'-DDT	7.018	6.178	191.6E6	1241.4E6	50.000	50.000
18)	B Endrin al...	6.912	6.253	139.5E6	827.0E6	50.000	50.000
19)	B Endosulfa...	7.147	6.477	193.3E6	1174.7E6	50.000	50.000
20)	A Methoxychlor	7.490	6.749	99144906	631.9E6	50.000	50.000
21)	B Endrin ke...	7.627	6.986	205.0E6	1295.3E6	50.000	50.000
22)	Mirex	8.110	7.179	151.5E6	980.0E6	50.000	50.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD081825\
Data File : PD089966.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2025 11:53
Operator : AR\AJ
Sample : PSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 19 02:23:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081825.M
Quant Title : GC Extractables
QLast Update : Tue Aug 19 02:15:07 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

