

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082021\
 Data File : PD065082.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Aug 2021 01:05
 Operator : AR\AJ
 Sample : PTOXCCC500
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PTOXCCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 20 09:00:26 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX082021.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 20 02:56:28 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds						
1) SA Tetrachlo...	3.496	4.013	50343246	382.3E6	48.763	47.937
7) SA Decachlor...	8.240	9.083	59332953	238.4E6	48.896	49.065
Target Compounds						
2) Toxaphene-1	5.749	6.406	4461756	18473555	504.344	503.874
3) Toxaphene-2	5.910	6.599	4703210	30976507	510.232	510.090
4) Toxaphene-3	6.352	7.182	10478111	53194449	502.045	495.768
5) Toxaphene-4	6.596	7.268	11413236	53269770	507.130	509.386
6) Toxaphene-5	7.083	7.686	13025400	41139835	496.893	486.946


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082021\
Data File : PD065082.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Aug 2021 01:05
Operator : AR\AJ
Sample : PTOXCCC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 20 09:00:26 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX082021.M
Quant Title : GC Extractables
QLast Update : Fri Aug 20 02:56:28 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.496 min
Delta R.T.: 0.000 min
Response: 50343246
Conc: 48.76 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXCCC500




#1 Tetrachloro-m-xylene

R.T.: 4.013 min
Delta R.T.: 0.000 min
Response: 382286112
Conc: 47.94 ng/ml




#2 Toxaphene-1

R.T.: 5.749 min
Delta R.T.: 0.000 min
Response: 4461756
Conc: 504.34 ng/ml



#2 Toxaphene-1


R.T.: 6.406 min
Delta R.T.: 0.000 min
Response: 18473555
Conc: 503.87 ng/ml



#3 Toxaphene-2


R.T.: 5.910 min
Delta R.T.: 0.000 min
Response: 4703210
Conc: 510.23 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXCCC500




#3 Toxaphene-2

R.T.: 6.599 min
Delta R.T.: 0.000 min
Response: 30976507
Conc: 510.09 ng/ml




#4 Toxaphene-3

R.T.: 6.352 min
Delta R.T.: 0.001 min
Response: 10478111
Conc: 502.04 ng/ml



#4 Toxaphene-3


R.T.: 7.182 min
Delta R.T.: 0.000 min
Response: 53194449
Conc: 495.77 ng/ml



#5 Toxaphene-4


R.T.: 6.596 min
Delta R.T.: 0.000 min
Response: 11413236
Conc: 507.13 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXCCC500




#5 Toxaphene-4

R.T.: 7.268 min
Delta R.T.: 0.000 min
Response: 53269770
Conc: 509.39 ng/ml




#6 Toxaphene-5

R.T.: 7.083 min
Delta R.T.: 0.000 min
Response: 13025400
Conc: 496.89 ng/ml



#6 Toxaphene-5


R.T.: 7.686 min
Delta R.T.: 0.000 min
Response: 41139835
Conc: 486.95 ng/ml



#7 Decachlorobiphenyl

R.T.: 8.240 min
Delta R.T.: 0.000 min
Response: 59332953
Conc: 48.90 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXCCC500



#7 Decachlorobiphenyl

R.T.: 9.083 min
Delta R.T.: 0.000 min
Response: 238437617
Conc: 49.06 ng/ml