

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082121\
 Data File : PD065141.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Aug 2021 23:45
 Operator : AR\AJ
 Sample : M3434-07MS
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 SWCS-11-1015MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 21 03:51:31 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082021.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 20 02:25:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.495	4.012	20162501	176.1E6	19.615	22.235
28)	SA Decachlor...	8.239	9.083	22788219	87619054	20.112	18.514
Target Compounds							
2)	A alpha-BHC	3.935	4.407	63997933	490.3E6	44.717	42.205
3)	MA gamma-BHC...	4.218	4.693	61251297	468.2E6	44.645	44.200
4)	MA Heptachlor	4.517	5.197	59985008	465.2E6	46.055	46.395
5)	MB Aldrin	4.765	5.502	55792351	410.5E6	46.524	45.448
6)	B beta-BHC	4.466	4.869	30336853	217.4E6	46.411	48.181
7)	B delta-BHC	4.671	5.086	32167581	246.5E6	24.331	25.384
8)	B Heptachlo...	5.206	5.892	53969837	351.8E6	47.493	47.157
9)	A Endosulfan I	5.545	6.250	51233801	329.1E6	48.030	46.730
10)	B gamma-Chl...	5.432	6.126	54440508	357.0E6	48.149	46.688
11)	B alpha-Chl...	5.490	6.198	53420863	345.9E6	47.712	46.463
12)	B 4,4'-DDE	5.653	6.355	52264675	342.5E6	49.590	47.694
13)	MA Dieldrin	5.788	6.508	55844933	349.4E6	49.761	48.155
14)	MA Endrin	6.046	6.727	50512913	292.2E6	48.135	47.481
15)	B Endosulfa...	6.318	6.940	50288690	287.9E6	50.295	49.102
16)	A 4,4'-DDD	6.168	6.849	45783670	283.6E6	48.408	48.229
17)	MA 4,4'-DDT	6.408	7.147	47547576	284.1E6	47.811	47.382
18)	B Endrin al...	6.486	7.064	44657093	254.0E6	53.967	54.458
19)	B Endosulfa...	6.698	7.287	48463003	264.3E6	45.638	46.081
20)	A Methoxychlor	6.954	7.606	28818210	150.5E6	48.892	49.729
21)	B Endrin ke...	7.192	7.760	64491640	329.7E6	50.587	50.726
22)	Mirex	7.377	8.209	49339615	223.0E6	48.386	50.076

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082121\
Data File : PD065141.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Aug 2021 23:45
Operator : AR\AJ
Sample : M3434-07MS
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
SWCS-11-1015MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 21 03:51:31 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082021.M
Quant Title : GC Extractables
QLast Update : Fri Aug 20 02:25:41 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

