

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082121\
 Data File : PD065165.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Aug 2021 05:43
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 21 06:24:57 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082021.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 20 02:25:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.495	4.012	52007493	403.5E6	50.596	50.943
28)	SA Decachlor...	8.239	9.083	60043047	244.2E6	52.993	51.592
Target Compounds							
2)	A alpha-BHC	3.935	4.407	73762436	584.8E6	51.540	50.338
3)	MA gamma-BHC...	4.218	4.694	68542300	533.3E6	49.960	50.341
4)	MA Heptachlor	4.516	5.198	65111569	494.4E6	49.991	49.313
5)	MB Aldrin	4.764	5.503	61545261	449.4E6	51.322	49.747
6)	B beta-BHC	4.466	4.870	33354922	239.2E6	52.164	53.415
7)	B delta-BHC	4.671	5.086	68358812	507.5E6	51.705	52.253
8)	B Heptachlo...	5.207	5.892	57621142	428.3E6	50.706	57.404
9)	A Endosulfan I	5.545	6.250	54119139	367.4E6	50.735	52.168
10)	B gamma-Chl...	5.431	6.126	57227515	420.7E6	50.613	55.018
11)	B alpha-Chl...	5.490	6.198	56435501	389.5E6	50.404	52.322
12)	B 4,4'-DDE	5.653	6.356	54742688	366.8E6	51.941	51.087
13)	MA Dieldrin	5.788	6.508	57231284	367.8E6	50.996	50.690
14)	MA Endrin	6.046	6.727	52329485	272.7E6	49.866	44.302
15)	B Endosulfa...	6.318	6.941	51441316	298.3E6	51.448	50.864
16)	A 4,4'-DDD	6.168	6.850	48852165	305.0E6	51.652	51.871
17)	MA 4,4'-DDT	6.408	7.147	50411705	307.6E6	50.691	51.292
18)	B Endrin al...	6.486	7.065	41411043	239.4E6	50.044	51.335
19)	B Endosulfa...	6.698	7.287	53902620	294.7E6	50.760	51.395
20)	A Methoxychlor	6.954	7.607	30419440	156.4E6	51.609	51.672
21)	B Endrin ke...	7.192	7.761	66535312	336.9E6	52.190	51.841
22)	Mirex	7.377	8.210	52142191	234.4E6	51.134	52.658

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082121\
Data File : PD065165.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Aug 2021 05:43
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 21 06:24:57 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082021.M
Quant Title : GC Extractables
QLast Update : Fri Aug 20 02:25:41 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

