

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082121\
 Data File : PD065100.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Aug 2021 12:18
 Operator : AR\AJ
 Sample : PTOXCCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PTOXCCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 21 02:03:19 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX082021.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 20 02:56:28 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds						
1) SA Tetrachlo...	3.496	4.012	50974938	393.6E6	49.375	49.350
7) SA Decachlor...	8.240	9.082	59460964	245.9E6	49.001	50.607
Target Compounds						
2) Toxaphene-1	5.748	6.406	4367642	18215680	493.706	496.840
3) Toxaphene-2	5.909	6.598	4562229	29740349	494.938	489.734
4) Toxaphene-3	6.351	7.181	10318357	53989359	494.391	503.176
5) Toxaphene-4	6.597	7.267	11065004	53899776	491.657	515.410
6) Toxaphene-5	7.082	7.685	12742312	43206435	486.094	511.407


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082121\
Data File : PD065100.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Aug 2021 12:18
Operator : AR\AJ
Sample : PTOXCCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 21 02:03:19 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX082021.M
Quant Title : GC Extractables
QLast Update : Fri Aug 20 02:56:28 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.496 min
Delta R.T.: 0.000 min
Response: 50974938
Conc: 49.38 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXCCC500




#1 Tetrachloro-m-xylene

R.T.: 4.012 min
Delta R.T.: 0.000 min
Response: 393556573
Conc: 49.35 ng/ml




#2 Toxaphene-1

R.T.: 5.748 min
Delta R.T.: 0.000 min
Response: 4367642
Conc: 493.71 ng/ml



#2 Toxaphene-1


R.T.: 6.406 min
Delta R.T.: -0.002 min
Response: 18215680
Conc: 496.84 ng/ml



#3 Toxaphene-2


R.T.: 5.909 min
Delta R.T.: -0.001 min
Response: 4562229
Conc: 494.94 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXCCC500




#3 Toxaphene-2

R.T.: 6.598 min
Delta R.T.: -0.001 min
Response: 29740349
Conc: 489.73 ng/ml




#4 Toxaphene-3

R.T.: 6.351 min
Delta R.T.: 0.000 min
Response: 10318357
Conc: 494.39 ng/ml



#4 Toxaphene-3


R.T.: 7.181 min
Delta R.T.: -0.001 min
Response: 53989359
Conc: 503.18 ng/ml



#5 Toxaphene-4


R.T.: 6.597 min
Delta R.T.: 0.000 min
Response: 11065004
Conc: 491.66 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXCCC500




#5 Toxaphene-4

R.T.: 7.267 min
Delta R.T.: -0.002 min
Response: 53899776
Conc: 515.41 ng/ml




#6 Toxaphene-5

R.T.: 7.082 min
Delta R.T.: -0.002 min
Response: 12742312
Conc: 486.09 ng/ml



#6 Toxaphene-5


R.T.: 7.685 min
Delta R.T.: -0.001 min
Response: 43206435
Conc: 511.41 ng/ml



#7 Decachlorobiphenyl

R.T.: 8.240 min
Delta R.T.: 0.000 min
Response: 59460964
Conc: 49.00 ng/ml

Instrument :
ECD_D
ClientSampleId :
PTOXCCC500



#7 Decachlorobiphenyl

R.T.: 9.082 min
Delta R.T.: 0.000 min
Response: 245931775
Conc: 50.61 ng/ml