

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082121\
 Data File : PD065199.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Aug 2021 19:26
 Operator : AR\AJ
 Sample : M3463-31
 Misc :
 ALS Vial : 78 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 SWCS-25-0204

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 23 12:08:36 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082021.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 20 02:25:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.496	4.013	18548932	141.2E6	18.046	17.832
28)	SA Decachlor...	8.239	9.084	16573845	61890677	14.628	13.078
Target Compounds							
13)	MA Dieldrin	5.786	6.508	1230648	2740556	1.097	0.378 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082121\
Data File : PD065199.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Aug 2021 19:26
Operator : AR\AJ
Sample : M3463-31
Misc :
ALS Vial : 78 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
SWCS-25-0204

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 23 12:08:36 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082021.M
Quant Title : GC Extractables
QLast Update : Fri Aug 20 02:25:41 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.496 min
Delta R.T.: 0.000 min
Response: 18548932
Conc: 18.05 ng/ml

Instrument :
ECD_D
ClientSampleId :
SWCS-25-0204




#1 Tetrachloro-m-xylene

R.T.: 4.013 min
Delta R.T.: 0.000 min
Response: 141238408
Conc: 17.83 ng/ml




#13 Dieldrin

R.T.: 5.786 min
Delta R.T.: -0.003 min
Response: 1230648
Conc: 1.10 ng/ml



#13 Dieldrin


R.T.: 6.508 min
Delta R.T.: 0.000 min
Response: 2740556
Conc: 0.38 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.239 min
Delta R.T.: 0.000 min
Response: 16573845
Conc: 14.63 ng/ml

Instrument :
ECD_D
ClientSampleId :
SWCS-25-0204



#28 Decachlorobiphenyl

R.T.: 9.084 min
Delta R.T.: 0.002 min
Response: 61890677
Conc: 13.08 ng/ml