

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082123\
 Data File : PD077294.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Aug 2023 21:31
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 21 22:29:43 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD080323.M
 Quant Title : GC Extractables
 QLast Update : Thu Aug 03 14:17:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.542	2.761	117.0E6	68921182	49.438	50.598
28)	SA Decachlor...	9.082	7.901	132.4E6	69024982	48.325	48.233
Target Compounds							
2)	A alpha-BHC	3.999	3.263	195.4E6	99651841	50.886	51.517
3)	MA gamma-BHC...	4.332	3.593	186.7E6	95727257	50.736	51.457
4)	MA Heptachlor	4.923	3.932	179.0E6	90581607	49.637	49.861
5)	MB Aldrin	5.266	4.212	184.9E6	92989117	50.044	50.220
6)	B beta-BHC	4.529	3.890	73717493	38136971	48.987	49.989
7)	B delta-BHC	4.778	4.119	182.6E6	93285917	52.331	52.430
8)	B Heptachlo...	5.694	4.714	161.3E6	82498987	48.706	49.334
9)	A Endosulfan I	6.081	5.084	146.8E6	76249666	48.290	48.879
10)	B gamma-Chl...	5.951	4.964	158.8E6	81571111	48.798	49.175
11)	B alpha-Chl...	6.030	5.028	158.9E6	81548393	48.461	48.832
12)	B 4,4'-DDE	6.202	5.217	144.9E6	74234956	48.850	51.191
13)	MA Dieldrin	6.356	5.349	158.4E6	83273992	48.390	49.433
14)	MA Endrin	6.587	5.624	125.2E6	68154140	45.047	46.578
15)	B Endosulfa...	6.807	5.919	131.5E6	70541130	41.342	47.855
16)	A 4,4'-DDD	6.720	5.772	117.3E6	61141377	49.864	54.980
17)	MA 4,4'-DDT	7.036	6.023	113.2E6	58541617	46.550	52.665
18)	B Endrin al...	6.937	6.098	101.6E6	55182517	46.527	48.474
19)	B Endosulfa...	7.172	6.321	121.6E6	66285845	46.683	49.700
20)	A Methoxychlor	7.512	6.598	63555249	31529336	44.856	52.730
21)	B Endrin ke...	7.658	6.827	146.0E6	81670392	48.467	51.511
22)	Mirex	8.133	7.008	105.0E6	61716469	46.025	47.665

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082123\
Data File : PD077294.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Aug 2023 21:31
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 21 22:29:43 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD080323.M
Quant Title : GC Extractables
QLast Update : Thu Aug 03 14:17:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

