

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082321\
 Data File : PD065247.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Aug 2021 17:34
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

mohammad
 8/24/2021 3:29:32 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 24 04:19:31 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082021.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 20 02:25:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.499	4.013	52182156	400.7E6	50.766	50.585
28)	SA Decachlor...	8.247	9.088	62823578	247.6E6	55.447	52.328
Target Compounds							
2)	A alpha-BHC	3.940	4.408	75477424	589.5E6	52.738	50.744
3)	MA gamma-BHC...	4.223	4.695	71063907	567.0E6	51.798	53.525
4)	MA Heptachlor	4.521	5.199	66119426	485.7E6	50.764	48.447
5)	MB Aldrin	4.770	5.504	64185305	452.6E6	53.523	50.108
6)	B beta-BHC	4.471	4.870	35493681	238.0E6	56.241	53.133
7)	B delta-BHC	4.676	5.088	71303501	500.4E6	53.933	51.524
8)	B Heptachlo...	5.212	5.894	60294477	385.3E6	53.058	51.646
9)	A Endosulfan I	5.551	6.252	57035893	357.9E6	53.470	50.815
10)	B gamma-Chl...	5.438	6.128	60565095	385.6E6	53.565	50.427
11)	B alpha-Chl...	5.496	6.200	59880545	377.7E6	53.481	50.733
12)	B 4,4'-DDE	5.659	6.358	57820667	367.3E6	54.862	51.150
13)	MA Dieldrin	5.794	6.510	60769589	374.5E6	54.149	51.607
14)	MA Endrin	6.052	6.729	55631340	296.7E6	53.013	48.209
15)	B Endosulfa...	6.324	6.943	54377560	303.4E6	54.385	51.734
16)	A 4,4'-DDD	6.174	6.851	52360056	311.8E6	55.361	53.031m
17)	MA 4,4'-DDT	6.415	7.150	52688001	304.0E6	52.980	50.691
18)	B Endrin al...	6.493	7.068	43707805	242.1E6	52.819	51.912
19)	B Endosulfa...	6.705	7.290	57580861	299.3E6	54.224	52.200
20)	A Methoxychlor	6.961	7.610	31681668	152.8E6	53.750	50.498
21)	B Endrin ke...	7.199	7.764	71419275	344.3E6	56.020	52.981
22)	Mirex	7.384	8.213	55407070	236.7E6	54.336	53.157

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082321\
Data File : PD065247.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Aug 2021 17:34
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

mohammad
8/24/2021 3:29:32 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 24 04:19:31 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082021.M
Quant Title : GC Extractables
QLast Update : Fri Aug 20 02:25:41 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

