

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082321\
 Data File : PD065253.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Aug 2021 19:03
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 24 04:21:42 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082021.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 20 02:25:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.496	4.013	54823894	382.9E6	53.336	48.342
28) SA Decachlor...	8.241	9.085	53669030	208.6E6	47.367	44.088
Target Compounds						
2) A alpha-BHC	3.935	4.408	77731484	569.8E6	54.313	49.047
3) MA gamma-BHC...	4.219	4.695	73233543	528.8E6	53.379	49.916
4) MA Heptachlor	4.517	5.198	67444286	475.1E6	51.782	47.383
5) MB Aldrin	4.765	5.504	64726127	435.9E6	53.974	48.255
6) B beta-BHC	4.467	4.871	36565558	238.1E6	58.284	53.146
7) B delta-BHC	4.672	5.087	72253426	505.1E6	54.651	52.007
8) B Heptachlo...	5.207	5.893	60951551	375.4E6	53.637	50.314
9) A Endosulfan I	5.546	6.251	57266650	335.2E6	53.686	47.600
10) B gamma-Chl...	5.432	6.127	60869982	363.6E6	53.835	47.555
11) B alpha-Chl...	5.491	6.199	60013502	354.0E6	53.600	47.558
12) B 4,4'-DDE	5.654	6.357	57833906	339.9E6	54.874	47.337
13) MA Dieldrin	5.789	6.509	60362761	338.4E6	53.787	46.630
14) MA Endrin	6.046	6.728	53925083	263.1E6	51.387	42.740
15) B Endosulfa...	6.319	6.941	52728720	269.9E6	52.736	46.026
16) A 4,4'-DDD	6.169	6.851	51011586	278.4E6	53.935	47.337
17) MA 4,4'-DDT	6.409	7.149	50064716	264.0E6	50.342	44.016
18) B Endrin al...	6.487	7.066	41688898	214.7E6	50.380	46.030
19) B Endosulfa...	6.699	7.288	54107806	262.2E6	50.954	45.727
20) A Methoxychlor	6.955	7.608	29415972	132.3E6	49.906	43.713
21) B Endrin ke...	7.193	7.762	64284185	290.5E6	50.424	44.695
22) Mirex	7.379	8.211	49479522	195.1E6	48.523	43.816

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082321\
 Data File : PD065253.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Aug 2021 19:03
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 24 04:21:42 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082021.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 20 02:25:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

