

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082323\
 Data File : PD077358.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Aug 2023 18:08
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 24 04:23:22 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD080323.M
 Quant Title : GC Extractables
 QLast Update : Thu Aug 03 14:17:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.550	2.761	128.3E6	75596584	54.217	55.499
28)	SA Decachlor...	9.091	7.904	142.1E6	72349718	51.844	50.556
Target Compounds							
2)	A alpha-BHC	4.007	3.263	216.9E6	110.1E6	56.500	56.897
3)	MA gamma-BHC...	4.340	3.593	205.8E6	105.2E6	55.925	56.549
4)	MA Heptachlor	4.931	3.933	196.5E6	100.5E6	54.499	55.322
5)	MB Aldrin	5.274	4.213	206.3E6	103.2E6	55.835	55.714
6)	B beta-BHC	4.538	3.890	78576846	41340490	52.216	54.188
7)	B delta-BHC	4.786	4.120	197.3E6	101.6E6	56.561	57.096
8)	B Heptachlo...	5.701	4.715	174.2E6	91049687	52.620	54.447
9)	A Endosulfan I	6.088	5.085	159.8E6	82619067	52.588	52.962
10)	B gamma-Chl...	5.958	4.965	174.0E6	89830729	53.466	54.154
11)	B alpha-Chl...	6.037	5.030	173.2E6	89594738	52.808	53.650
12)	B 4,4'-DDE	6.210	5.219	158.2E6	81018423	53.317	55.868
13)	MA Dieldrin	6.363	5.350	174.5E6	91717687	53.315	54.446
14)	MA Endrin	6.594	5.626	139.2E6	78659202	50.109	53.757
15)	B Endosulfa...	6.814	5.920	143.1E6	76473429	44.983	51.880
16)	A 4,4'-DDD	6.729	5.773	128.2E6	65217364	54.498	58.645
17)	MA 4,4'-DDT	7.043	6.024	122.4E6	62642645	50.349	56.354
18)	B Endrin al...	6.944	6.100	109.9E6	59335420	50.314	52.122
19)	B Endosulfa...	7.179	6.322	132.1E6	70385712	50.712	52.774
20)	A Methoxychlor	7.519	6.600	68466634	33552136	48.322	56.113
21)	B Endrin ke...	7.665	6.830	154.9E6	85380223	51.425	53.851
22)	Mirex	8.140	7.011	112.4E6	65804756	49.262	50.823

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082323\
Data File : PD077358.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Aug 2023 18:08
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 24 04:23:22 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD080323.M
Quant Title : GC Extractables
QLast Update : Thu Aug 03 14:17:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

