

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082423\
 Data File : PD077359.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Aug 2023 08:44
 Operator : AR\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 25 01:10:39 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD080323.M
 Quant Title : GC Extractables
 QLast Update : Thu Aug 03 14:17:41 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds						
2) A	alpha-BHC	0.000	3.287f	0	130055	N.D. 0.067 #
5) MB	Aldrin	5.271	0.000	238234	0	0.064 N.D. #
7) B	delta-BHC	4.779	0.000	1265651	0	0.363 N.D. #
8) B	Heptachlo...	5.698	0.000	3485598	0	1.053 N.D. #
14) MA	Endrin	0.000	5.644f	0	4539346	N.D. 3.102 #
17) MA	4,4'-DDT	7.053f	0.000	582522	0	0.240 N.D. #
24)	Chlordane-2	0.000	4.340	0	6201732	N.D. 110.281 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082423\
Data File : PD077359.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Aug 2023 08:44
Operator : AR\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 25 01:10:39 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD080323.M
Quant Title : GC Extractables
QLast Update : Thu Aug 03 14:17:41 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T. : 3.543 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 2.762 min
Response: 0
Conc: N.D.




#2 alpha-BHC

R.T.: 0.000 min
Exp R.T. : 4.000 min
Response: 0
Conc: N.D.



#2 alpha-BHC


R.T.: 3.287 min
Delta R.T.: 0.024 min
Response: 130055
Conc: 0.07 ng/ml



#5 Aldrin


R.T.: 5.271 min
Delta R.T.: 0.003 min
Response: 238234
Conc: 0.06 ng/ml

Instrument :
ECD_D
ClientSampleId :




#5 Aldrin

R.T.: 0.000 min
Exp R.T. : 4.213 min
Response: 0
Conc: N.D.




#7 delta-BHC

R.T.: 4.779 min
Delta R.T.: 0.000 min
Response: 1265651
Conc: 0.36 ng/ml



#7 delta-BHC


R.T.: 0.000 min
Exp R.T. : 4.120 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 5.698 min
Delta R.T.: 0.004 min
Response: 3485598
Conc: 1.05 ng/ml

Instrument :
ECD_D
ClientSampleId :




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 4.716 min
Response: 0
Conc: N.D.




#14 Endrin

R.T.: 0.000 min
Exp R.T. : 6.588 min
Response: 0
Conc: N.D.



#14 Endrin


R.T.: 5.644 min
Delta R.T.: 0.018 min
Response: 4539346
Conc: 3.10 ng/ml



#15 Endosulfan II


R.T.: 6.807 min
Delta R.T.: -0.001 min
Response: 226691
Conc: 0.07 ng/ml

Instrument :
ECD_D
ClientSampleId :




#15 Endosulfan II

R.T.: 5.901 min
Delta R.T.: -0.019 min
Response: -158447
Conc: N.D.




#17 4,4'-DDT

R.T.: 7.053 min
Delta R.T.: 0.016 min
Response: 582522
Conc: 0.24 ng/ml



#17 4,4'-DDT


R.T.: 0.000 min
Exp R.T.: 6.025 min
Response: 0
Conc: N.D.



#24 Chlordane-2


R.T.: 0.000 min
Exp R.T.: 5.240 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#24 Chlordane-2

R.T.: 4.340 min
Delta R.T.: 0.004 min
Response: 6201732
Conc: 110.28 ng/ml



#28 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T.: 9.086 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T.: 7.904 min
Response: 0
Conc: N.D.