

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082521\
 Data File : PD065285.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Aug 2021 22:30
 Operator : AR\AJ
 Sample : PTOXICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PTOXICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 25 05:49:14 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX082521.M
 Quant Title : GC Extractables
 QLast Update : Wed Aug 25 05:48:48 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds						
1) SA Tetrachlo...	3.495	4.012	100.6E6	634.2E6	96.357	93.968
7) SA Decachlor...	8.239	9.082	122.1E6	411.0E6	93.316	93.679
Target Compounds						
2) Toxaphene-1	5.748	6.406	8739840	29762376	899.908	965.298
3) Toxaphene-2	5.909	6.598	9368386	50705167	912.408	944.968
4) Toxaphene-3	6.351	7.181	20860406	89709402	917.001	933.719
5) Toxaphene-4	6.595	7.267	22767284	89693872	920.955	945.840
6) Toxaphene-5	7.083	7.685	25987553	74960868	915.593	971.400

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082521\
Data File : PD065285.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Aug 2021 22:30
Operator : AR\AJ
Sample : PTOXICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 25 05:49:14 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX082521.M
Quant Title : GC Extractables
QLast Update : Wed Aug 25 05:48:48 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Instrument :
ECD_D
ClientSample :
PTOXICC1000