

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082521\
 Data File : PD065318.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Aug 2021 18:04
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

mohammad
 8/26/2021 2:27:05 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 26 03:41:28 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082521.M
 Quant Title : GC Extractables
 QLast Update : Wed Aug 25 05:44:31 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.495	4.013	50244129	342.8E6	46.602	50.139
28)	SA Decachlor...	8.239	9.083	65127589	228.7E6	47.997	52.819
Target Compounds							
2)	A alpha-BHC	3.935	4.408	72336357	497.8E6	47.064	50.305
3)	MA gamma-BHC...	4.218	4.694	68953680	458.6E6	45.402	50.955
4)	MA Heptachlor	4.516	5.198	65886765	426.9E6	46.247	50.450
5)	MB Aldrin	4.764	5.503	63207327	397.3E6	47.839	51.057
6)	B beta-BHC	4.466	4.869	33914940	211.1E6	50.624	55.197
7)	B delta-BHC	4.671	5.085	68657250	445.0E6	46.652	51.364m
8)	B Heptachlo...	5.206	5.892	59942801	374.4E6	47.424	58.828
9)	A Endosulfan I	5.545	6.250	56804358	317.0E6	47.464	51.947
10)	B gamma-Chl...	5.431	6.126	60409541	338.7E6	47.464	51.470
11)	B alpha-Chl...	5.490	6.198	59647012	333.4E6	47.222	51.622
12)	B 4,4'-DDE	5.652	6.356	58535919	324.4E6	48.990	52.285
13)	MA Dieldrin	5.788	6.508	60421344	332.1E6	47.738	52.424
14)	MA Endrin	6.045	6.726	56519242	253.1E6	47.174	47.130
15)	B Endosulfa...	6.317	6.940	54012970	272.7E6	46.714	52.213
16)	A 4,4'-DDD	6.167	6.850	51761760	276.1E6	48.419	52.911
17)	MA 4,4'-DDT	6.408	7.147	54172852	277.1E6	46.773	51.529
18)	B Endrin al...	6.486	7.064	43507603	217.3E6	45.590	51.383
19)	B Endosulfa...	6.697	7.287	57213967	270.9E6	46.341	51.753
20)	A Methoxychlor	6.953	7.606	33078322	145.1E6	47.093	51.956
21)	B Endrin ke...	7.192	7.760	70072258	314.2E6	47.748	53.294
22)	Mirex	7.376	8.209	56066086	219.8E6	53.198	53.235

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082521\
Data File : PD065318.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Aug 2021 18:04
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

mohammad
8/26/2021 2:27:05 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 26 03:41:28 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082521.M
Quant Title : GC Extractables
QLast Update : Wed Aug 25 05:44:31 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

