

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082622\
 Data File : PD071663.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Aug 2022 16:43
 Operator : AR\AJ
 Sample : PEM304
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PEM094

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 08/29/2022
 Supervised By : Ankita Jodhani 08/29/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 27 07:23:02 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082622CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Aug 27 07:11:22 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	2.805	3.563	51827625	655.9E6	19.873	20.017
27)	SA Decachlor...	7.997	9.088	54422196	202.1E6	20.738	21.140
Target Compounds							
2)	A alpha-BHC	3.313	4.029	38643960	466.5E6	10.019	10.573
3)	MA gamma-BHC...	3.647	4.364	36382973	404.6E6	9.855	10.878
6)	B beta-BHC	3.944	4.571	16780669	159.4E6	10.564m	11.140
12)	B 4,4'-DDE	5.294	6.220	648487	3389341	0.211	0.232
14)	MA Endrin	5.702	6.613	141.6E6	590.6E6	47.273	44.890
16)	A 4,4'-DDD	5.850	6.746	3043772	14717835	1.156m	1.203
17)	MA 4,4'-DDT	6.103	7.056	287.2E6	1143.5E6	105.499	94.770
18)	B Endrin al...	6.178	6.971	4571333	19457133	1.976	1.907
20)	A Methoxychlor	6.683	7.533	366.9E6	1364.5E6	255.477	248.686
21)	B Endrin ke...	6.911	7.695	7445492	32870949	2.488	2.655

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082622\
Data File : PD071663.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Aug 2022 16:43
Operator : AR\AJ
Sample : PEM304
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 27 07:23:02 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082622CLP.M
Quant Title : GC Extractables
QLast Update : Sat Aug 27 07:11:22 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Reviewed By: Abdul Mirza 08/29/2022
Supervised By: Ankita Jodhani 08/29/2022

Manual Integrations
APPROVED

Instrument :
ECD_D
Client Sampled :
PEM094