

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082622\
 Data File : PD071690.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Aug 2022 22:49
 Operator : AR\AJ
 Sample : N4320-08MSD
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 C0AX9MSD

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 08/29/2022
 Supervised By :Ankita Jodhani 08/29/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 27 07:26:20 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082622CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Aug 27 07:11:22 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.805	3.563	43169255	548.3E6	16.553	16.732
27)	SA Decachlor...	7.996	9.088	87569847	323.6E6	33.370	33.851
Target Compounds							
2)	A alpha-BHC	3.314	4.029	221.0E6	2241.9E6	57.295	50.813
3)	MA gamma-BHC...	3.647	4.364	208.0E6	1836.7E6	56.328	49.385
4)	MA Heptachlor	3.992	4.942	198.2E6	1214.8E6	52.085m	44.990
5)	MB Aldrin	4.277	5.284	192.9E6	1012.5E6	54.196	46.868
6)	B beta-BHC	3.944	4.571	82261468	701.3E6	51.786m	49.000
7)	B delta-BHC	4.179	4.818	208.0E6	1389.8E6	59.532	52.877
8)	B Heptachlo...	4.784	5.716	189.1E6	873.1E6	58.029	51.629
9)	A Endosulfan I	5.157	6.102	173.8E6	766.5E6	55.150	48.804m
10)	B trans-Chl...	5.036	5.970	188.2E6	827.9E6	55.592	48.906
11)	B cis-Chlor...	5.101	6.046	185.9E6	815.6E6	55.723	50.038m
12)	B 4,4'-DDE	5.292	6.221	335.9E6	1417.0E6	109.517	97.109
13)	MA Dieldrin	5.424	6.381	396.7E6	1650.3E6	114.567	103.118
14)	MA Endrin	5.701	6.613	348.0E6	1362.4E6	116.189	103.549
15)	B Endosulfa...	5.996	6.839	322.6E6	1320.2E6	121.364	109.524
16)	A 4,4'-DDD	5.850	6.745	299.1E6	1242.5E6	113.628	101.542
17)	MA 4,4'-DDT	6.104	7.056	316.1E6	1278.9E6	116.132	105.994
18)	B Endrin al...	6.177	6.970	255.5E6	1048.4E6	110.465	102.765
19)	B Endosulfa...	6.402	7.203	302.4E6	1197.7E6	113.502	105.461
20)	A Methoxychlor	6.683	7.532	778.8E6	2847.2E6	542.235	518.905
21)	B Endrin ke...	6.911	7.694	368.2E6	1381.3E6	123.048	111.590

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082622\
Data File : PD071690.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Aug 2022 22:49
Operator : AR\AJ
Sample : N4320-08MSD
Misc :
ALS Vial : 29 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 27 07:26:20 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082622CLP.M
Quant Title : GC Extractables
QLast Update : Sat Aug 27 07:11:22 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

