

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD D\Data\PD082818\
 Data File : PD049045.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Aug 2018 13:36
 Operator : AJ\SJ
 Sample : INDA313
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA313

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 29 02:03:39 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081618CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Aug 16 13:32:25 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.370	3.967	22820751	349.8E6	18.697	19.893
27)	SA Decachlor...	8.065	8.992	54158943	615.0E6	37.613	37.991
Target Compounds							
2)	A alpha-BHC	3.800	4.348	35991601	534.9E6	18.626	19.702
3)	MA gamma-BHC...	4.078	4.628	33782995	496.8E6	18.176	19.376
4)	MA Heptachlor	4.366	5.141	34035522	484.2E6	18.973	19.613
9)	A Endosulfan I	5.380	6.179	29081471	407.0E6	18.691	19.001
13)	MA Dieldrin	5.620	6.433	65204294	839.7E6	37.188	37.867
14)	MA Endrin	5.876	6.648	56505884	654.9E6	35.968	36.778
16)	A 4,4'-DDD	6.007	6.767	52993634	568.2E6	37.166	37.472
17)	MA 4,4'-DDT	6.244	7.066	54381287	587.4E6	36.834	36.432
20)	A Methoxychlor	6.793	7.520	144.0E6	1240.4E6	184.804	181.193

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD D\Data\PD082818\
Data File : PD049045.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Aug 2018 13:36
Operator : AJ\SJ
Sample : INDA313
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampled :
INDA313

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 29 02:03:39 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081618CLP.M
Quant Title : GC Extractables
QLast Update : Thu Aug 16 13:32:25 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Ini. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

