

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082825\
 Data File : PD090055.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Aug 2025 10:50
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 28 23:24:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081825.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 19 13:16:54 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.549	2.877	143.5E6	998.8E6	48.902	45.298
28)	SA Decachlor...	9.068	8.064	211.0E6	1188.3E6	48.889	45.905
Target Compounds							
2)	A alpha-BHC	3.998	3.389	320.4E6	1609.0E6	53.377	47.410
3)	MA gamma-BHC...	4.328	3.725	302.0E6	1486.4E6	52.739	47.687
4)	MA Heptachlor	4.927	4.077	291.4E6	1420.3E6	51.947	45.535
5)	MB Aldrin	5.268	4.363	286.2E6	1449.6E6	52.858	47.597
6)	B beta-BHC	4.515	4.021	113.6E6	628.4E6	50.642	47.139
7)	B delta-BHC	4.763	4.257	296.5E6	1497.0E6	57.273	47.448
8)	B Heptachlo...	5.688	4.867	257.5E6	1323.0E6	52.481	47.771
9)	A Endosulfan I	6.071	5.241	243.4E6	1245.7E6	52.741	47.506
10)	B gamma-Chl...	5.944	5.120	260.9E6	1414.7E6	53.349	47.777
11)	B alpha-Chl...	6.024	5.184	258.3E6	1346.7E6	52.474	47.188
12)	B 4,4'-DDE	6.193	5.369	232.6E6	1363.1E6	52.999	47.292
13)	MA Dieldrin	6.344	5.506	264.6E6	1395.7E6	54.037	47.656
14)	MA Endrin	6.571	5.783	219.2E6	1237.5E6	52.903	46.468
15)	B Endosulfa...	6.783	6.074	223.5E6	1209.4E6	52.962	47.814
16)	A 4,4'-DDD	6.702	5.923	188.1E6	1144.4E6	54.391	47.598
17)	MA 4,4'-DDT	7.018	6.176	189.7E6	1117.7E6	49.791	44.586
18)	B Endrin al...	6.912	6.252	159.4E6	880.7E6	55.598	51.142
19)	B Endosulfa...	7.147	6.475	208.8E6	1174.5E6	53.279	48.175
20)	A Methoxychlor	7.490	6.747	96219946	553.7E6	47.975	42.938
21)	B Endrin ke...	7.627	6.984	224.8E6	1299.0E6	54.246	48.751
22)	Mirex	8.110	7.178	160.5E6	968.5E6	50.883	46.887

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD082825\
Data File : PD090055.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Aug 2025 10:50
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 28 23:24:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081825.M
Quant Title : GC Extractables
QLast Update : Tue Aug 19 13:16:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

