

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD083024\
 Data File : PD084807.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Aug 2024 14:46
 Operator : AR\AJ
 Sample : P3721-02
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 DCXY6

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 29 09:00:56 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 27 02:47:48 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.537	2.905	11052204	41805067	10.928	11.813
27)	SA Decachlor...	9.035	8.109	12257073	43353811	9.385	12.468 #
Target Compounds							
8)	B Heptachlo...	0.000	4.849f	0	1718270	N.D.	0.406
18)	B Endrin al...	6.831f	0.000	5317019	0	5.887	N.D. #
19)	B Endosulfa...	7.173f	0.000	1496111	0	1.299	N.D. #
20)	A Methoxychlor	7.420f	0.000	850025	0	1.632	N.D. #
21)	B Endrin ke...	7.593	0.000	6970311	0	5.701	N.D. #
22)	Toxaphene-1	5.780f	0.000	1049770	0	249.063	N.D. #
23)	Toxaphene-2	6.462f	0.000	3259390	0	221.601	N.D. #
25)	Toxaphene-4	7.173f	0.000	1496111	0	58.144	N.D. #
26)	Toxaphene-5	7.893	7.339	9403154	22494778	499.128	546.862


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD083024\
Data File : PD084807.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Aug 2024 14:46
Operator : AR\AJ
Sample : P3721-02
Misc :
ALS Vial : 19 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 29 09:00:56 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 27 02:47:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.537 min
Delta R.T.: 0.000 min
Response: 11052204
Conc: 10.93 ng/ml

Instrument :
ECD_D
ClientSampleId :
DCXY6




#1 Tetrachloro-m-xylene

R.T.: 2.905 min
Delta R.T.: 0.000 min
Response: 41805067
Conc: 11.81 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 5.670 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 4.849 min
Delta R.T.: -0.049 min
Response: 1718270
Conc: 0.41 ng/ml



#18 Endrin aldehyde


R.T.: 6.831 min
Delta R.T.: -0.059 min
Response: 5317019
Conc: 5.89 ng/ml

Instrument :
ECD_D
ClientSampleId :
DCXY6




#18 Endrin aldehyde

R.T.: 0.000 min
Exp R.T. : 6.289 min
Response: 0
Conc: N.D.




#19 Endosulfan Sulfate

R.T.: 7.173 min
Delta R.T.: 0.050 min
Response: 1496111
Conc: 1.30 ng/ml



#19 Endosulfan Sulfate


R.T.: 0.000 min
Exp R.T. : 6.513 min
Response: 0
Conc: N.D.



#20 Methoxychlor


R.T.: 7.420 min
Delta R.T.: -0.047 min
Response: 850025
Conc: 1.63 ng/ml

Instrument :
ECD_D
ClientSampleId :
DCXY6




#20 Methoxychlor

R.T.: 0.000 min
Exp R.T. : 6.785 min
Response: 0
Conc: N.D.




#21 Endrin ketone

R.T.: 7.593 min
Delta R.T.: -0.010 min
Response: 6970311
Conc: 5.70 ng/ml



#21 Endrin ketone


R.T.: 0.000 min
Exp R.T. : 7.023 min
Response: 0
Conc: N.D.



#22 Toxaphene-1


R.T.: 5.780 min
Delta R.T.: -0.051 min
Response: 1049770
Conc: 249.06 ng/ml

Instrument :
ECD_D
ClientSampleId :
DCXY6




#22 Toxaphene-1

R.T.: 0.000 min
Exp R.T. : 5.173 min
Response: 0
Conc: N.D.




#23 Toxaphene-2

R.T.: 6.462 min
Delta R.T.: 0.044 min
Response: 3259390
Conc: 221.60 ng/ml



#23 Toxaphene-2


R.T.: 0.000 min
Exp R.T. : 5.502 min
Response: 0
Conc: N.D.



#25 Toxaphene-4


R.T.: 7.173 min
Delta R.T.: 0.048 min
Response: 1496111
Conc: 58.14 ng/ml

Instrument :
ECD_D
ClientSampleId :
DCXY6




#25 Toxaphene-4

R.T.: 0.000 min
Exp R.T.: 7.232 min
Response: 0
Conc: N.D.




#26 Toxaphene-5

R.T.: 7.893 min
Delta R.T.: -0.009 min
Response: 9403154
Conc: 499.13 ng/ml



#26 Toxaphene-5


R.T.: 7.339 min
Delta R.T.: -0.024 min
Response: 22494778
Conc: 546.86 ng/ml



#27 Decachlorobiphenyl

R.T.: 9.035 min
Delta R.T.: 0.003 min
Response: 12257073
Conc: 9.38 ng/ml

Instrument :
ECD_D
ClientSampleId :
DCXY6



#27 Decachlorobiphenyl

R.T.: 8.109 min
Delta R.T.: 0.001 min
Response: 43353811
Conc: 12.47 ng/ml