

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD090123\
 Data File : PD077522.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Aug 2023 11:14
 Operator : AR\AJ
 Sample : 04175-19
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 EZYD2

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/01/2023
 Supervised By : Ankita Jodhani 09/05/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 01 02:01:58 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081423CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 15 01:53:32 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.542	2.762	25026231	19376778	11.399	15.554 #
27)	SA Decachlor...	9.086	7.902	44964754	21856897	17.671	17.534
Target Compounds							
5)	MB Aldrin	5.261	4.221	4928099	1557766	1.422	0.900 #
7)	B delta-BHC	4.777	4.127	3652440	6001398	1.102	3.455 #
8)	B Heptachlo...	5.689	4.711	2446935	4472966	0.767	2.734m#
12)	B 4,4'-DDE	6.217	5.234	8503087	3540335	2.969	2.601m
17)	MA 4,4'-DDT	7.007	6.018	4577431	12968952	2.075m	12.526 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD090123\
Data File : PD077522.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Aug 2023 11:14
Operator : AR\AJ
Sample : 04175-19
Misc :
ALS Vial : 12 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 01 02:01:58 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081423CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 15 01:53:32 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integrations
APPROVED
Reviewed By: Abdul Mirza 09/01/2023
Supervised By: Ankita Jodhani 09/05/2023

Instrument :
ECD_D
ClientSampled :
EZYD2