

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD090222\
 Data File : PD071721.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Sep 2022 17:32
 Operator : AR\AJ
 Sample : PIBLK331
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 03 02:06:32 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082622CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Aug 27 07:11:22 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.805	3.563	56281271	722.9E6	21.581	22.062
27)	SA Decachlor...	7.993	9.086	107.2E6	402.3E6	40.860	42.084
Target Compounds							
2)	A alpha-BHC	3.302	4.034	185852	-72249	<MDL	N.D. #
3)	MA gamma-BHC...	3.652	4.365	161371	4179930	<MDL	0.112 #
6)	B beta-BHC	3.906f	4.586	207929	5844502	0.131	0.408 #
8)	B Heptachlo...	0.000	5.713	0	6303774	N.D.	0.373
9)	A Endosulfan I	0.000	6.089	0	716200	N.D.	<MDL
11)	B cis-Chlor...	0.000	6.057	0	273970	N.D.	<MDL
12)	B 4,4'-DDE	0.000	6.227	0	141149	N.D.	<MDL
13)	MA Dieldrin	0.000	6.370	0	808893	N.D.	<MDL
14)	MA Endrin	5.715	6.680f	466652	7346284	0.156	0.558 #
15)	B Endosulfa...	0.000	6.796f	0	1016966	N.D.	<MDL
16)	A 4,4'-DDD	0.000	6.736	0	532239	N.D.	<MDL
18)	B Endrin al...	0.000	7.000	0	101036	N.D.	<MDL
19)	B Endosulfa...	0.000	7.221	0	137569	N.D.	<MDL
21)	B Endrin ke...	6.941	7.628f	555533	1030618	0.186	<MDL #
22)	Toxaphene-1	0.000	5.886	0	401656	N.D.	6.765
23)	Toxaphene-2	5.715f	6.277	466652	142093	25.350	1.195 #
25)	Toxaphene-4	6.829	7.221	144358	137569	1.645	0.524 #
26)	Toxaphene-5	0.000	7.628	0	1030618	N.D.	6.910

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD090222\
Data File : PD071721.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Sep 2022 17:32
Operator : AR\AJ
Sample : PIBLK331
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 03 02:06:32 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082622CLP.M
Quant Title : GC Extractables
QLast Update : Sat Aug 27 07:11:22 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

