

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD090419CLP.M
 Title : GC Extractables
 Last Update : Thu Sep 05 06:52:51 2019
 Response Via : Initial Calibration

Calibration Files

3 =PD054400.D 1 =PD054398.D 2 =PD054399.D
 4 =PD054401.D 5 =PD054402.D

Compound		3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene	7.414	7.264	7.932	7.479	7.363	7.490	E5	3.46
2) A	alpha-BHC						0.000		-1.00
3) MA	gamma-BHC (Lindane)						0.000		-1.00
4) MA	Heptachlor						0.000		-1.00
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I						0.000		-1.00
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin						0.000		-1.00
14) MA	Endrin						0.000		-1.00
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD						0.000		-1.00
17) MA	4,4'-DDT						0.000		-1.00
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor						0.000		-1.00
21) B	Endrin ketone						0.000		-1.00
22)	Toxaphene-1	5.456	5.264	5.392	5.084	5.057	5.251	E3	3.40
23)	Toxaphene-2	5.231	5.355	5.531	5.200	5.258	5.315	E3	2.52
24)	Toxaphene-3	4.564	4.781	4.444	4.694	4.630	4.623	E3	2.77
25)	Toxaphene-4	1.387	1.335	1.375	1.417	1.359	1.375	E4	2.24
26)	Toxaphene-5	1.639	1.705	1.631	1.639	1.591	1.641	E4	2.51
27) SA	Decachlorobiphenyl	7.410	8.988	7.667	7.207	6.668	7.588	E5	11.39

Signal #2 Calibration Files

3 =PD054400.D 1 =PD054398.D 2 =PD054399.D
 4 =PD054401.D 5 =PD054402.D

Compound			3	1	2	4	5	Avg		%RSD
1) SA	Tetrachloro-m-xylene		0.957	1.039	1.057	1.022	0.994	1.014	E6	3.89
2) A	alpha-BHC							0.000		-1.00
3) MA	gamma-BHC (Lindane)							0.000		-1.00
4) MA	Heptachlor							0.000		-1.00
5) MB	Aldrin							0.000		-1.00
6) B	beta-BHC							0.000		-1.00
7) B	delta-BHC							0.000		-1.00
8) B	Heptachlor epoxide							0.000		-1.00
9) A	Endosulfan I							0.000		-1.00
10) B	trans-Chlordane							0.000		-1.00
11) B	cis-Chlordane							0.000		-1.00
12) B	4,4'-DDE							0.000		-1.00
13) MA	Dieldrin							0.000		-1.00
14) MA	Endrin							0.000		-1.00
15) B	Endosulfan II							0.000		-1.00
16) A	4,4'-DDD							0.000		-1.00
17) MA	4,4'-DDT							0.000		-1.00
18) B	Endrin aldehyde							0.000		-1.00
19) B	Endosulfan Sulfate							0.000		-1.00
20) A	Methoxychlor							0.000		-1.00
21) B	Endrin ketone							0.000		-1.00
22)	Toxaphene-1		6.135	6.439	6.826	6.406	5.932	6.348	E3	5.33
23)	Toxaphene-2		7.846	8.537	8.163	8.488	7.962	8.199	E3	3.76

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD090419CLP.M
 Title : GC Extractables
 Last Update : Thu Sep 05 06:52:51 2019
 Response Via : Initial Calibration

Calibration Files

3 =PD054400.D 1 =PD054398.D 2 =PD054399.D
 4 =PD054401.D 5 =PD054402.D

	Compound	3	1	2	4	5	Avg		%RSD
24)	Toxaphene-3	1.051	1.189	1.167	1.067	0.989	1.092	E4	7.63
25)	Toxaphene-4	3.092	2.870	3.032	3.023	2.975	2.999	E4	2.77
26)	Toxaphene-5	1.714	1.757	1.762	1.716	1.659	1.722	E4	2.40
27) SA	Decachlorobiphenyl	0.985	1.180	1.103	0.963	0.901	1.026	E6	11.01

(#) = Out of Range