

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
 Method File : PD090419CLP.M
 Title : GC Extractables
 Last Update : Thu Sep 05 07:28:44 2019
 Response Via : Initial Calibration

Calibration Files

3 =PD054407.D 1 =PD054403.D 2 =PD054405.D
 4 =PD054409.D 5 =PD054411.D

Compound		3	1	2	4	5	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	6.958	7.292	7.102	7.222	6.766	7.068	E5	2.99
2) A	alpha-BHC	1.064	1.108	1.107	1.104	1.050	1.087	E6	2.52
3) MA	gamma-BHC (Lindane)	0.975	0.980	1.001	1.021	0.956	0.987	E6	2.55
4) MA	Heptachlor	7.367	7.611	7.507	7.673	7.243	7.480	E5	2.36
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I	8.549	9.172	8.420	8.515	8.009	8.533	E5	4.89
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin	9.375	9.257	9.184	9.719	9.174	9.342	E5	2.41
14) MA	Endrin	7.641	8.433	7.447	7.813	7.065	7.680	E5	6.57
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD	6.867	7.252	6.815	6.998	6.677	6.922	E5	3.14
17) MA	4,4'-DDT	6.240	6.030	5.987	6.620	6.582	6.292	E5	4.74
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor	2.806	2.855	2.786	2.847	2.723	2.803	E5	1.90
21) B	Endrin ketone						0.000		-1.00
22)	Toxaphene-1						0.000		-1.00
23)	Toxaphene-2						0.000		-1.00
24)	Toxaphene-3						0.000		-1.00
25)	Toxaphene-4						0.000		-1.00
26)	Toxaphene-5						0.000		-1.00
27) SA	Decachlorobiphenyl	6.663	7.205	6.936	6.605	6.429	6.768	E5	4.50

Signal #2 Calibration Files

3 =PD054407.D 1 =PD054403.D 2 =PD054405.D
 4 =PD054409.D 5 =PD054411.D

Compound		3	1	2	4	5	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	0.985	1.028	0.984	1.000	0.938	0.987	E6	3.32
2) A	alpha-BHC	1.438	1.561	1.473	1.523	1.461	1.491	E6	3.35
3) MA	gamma-BHC (Lindane)	1.378	1.414	1.400	1.442	1.372	1.401	E6	2.03
4) MA	Heptachlor	1.337	1.354	1.418	1.435	1.349	1.379	E6	3.23
5) MB	Aldrin						0.000		-1.00
6) B	beta-BHC						0.000		-1.00
7) B	delta-BHC						0.000		-1.00
8) B	Heptachlor epoxide						0.000		-1.00
9) A	Endosulfan I	1.201	1.350	1.197	1.276	1.190	1.243	E6	5.59
10) B	trans-Chlordane						0.000		-1.00
11) B	cis-Chlordane						0.000		-1.00
12) B	4,4'-DDE						0.000		-1.00
13) MA	Dieldrin	1.332	1.337	1.319	1.403	1.348	1.348	E6	2.42
14) MA	Endrin	1.010	1.055	0.990	1.065	0.989	1.022	E6	3.53
15) B	Endosulfan II						0.000		-1.00
16) A	4,4'-DDD	1.056	1.036	1.043	1.107	1.087	1.066	E6	2.82
17) MA	4,4'-DDT	1.005	1.035	0.985	1.096	1.074	1.039	E6	4.45
18) B	Endrin aldehyde						0.000		-1.00
19) B	Endosulfan Sulfate						0.000		-1.00
20) A	Methoxychlor	5.278	5.243	5.240	5.358	5.104	5.245	E5	1.75
21) B	Endrin ketone						0.000		-1.00
22)	Toxaphene-1						0.000		-1.00
23)	Toxaphene-2						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\
Method File : PD090419CLP.M
Title : GC Extractables
Last Update : Thu Sep 05 07:28:44 2019
Response Via : Initial Calibration

Calibration Files

3	=PD054407.D	1	=PD054403.D	2	=PD054405.D
4	=PD054409.D	5	=PD054411.D		

	Compound	3	1	2	4	5	Avg	%RSD
24)	Toxaphene-3						0.000	-1.00
25)	Toxaphene-4						0.000	-1.00
26)	Toxaphene-5						0.000	-1.00
27) SA	Decachlorobiphenyl	9.271	9.829	9.587	9.254	8.834	9.355 E5	4.02

(#) = Out of Range