

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD090421\
 Data File : PD065569.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Sep 2021 21:52
 Operator : AR\AJ
 Sample : PTOXICV500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 04 01:38:41 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\DTX090421.M
 Quant Title : GC Extractables
 QLast Update : Sat Sep 04 01:38:06 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA	Tetrachloro-m-xylene	968.818	975.345 E3	-0.7	99	0.00
2	Toxaphene-1	10.756	11.004 E3	-2.3	101	0.00
3	Toxaphene-2	12.043	11.702 E3	2.8	98	0.00
4	Toxaphene-3	26.261	26.434 E3	-0.7	98	0.00
5	Toxaphene-4	29.092	28.647 E3	1.5	97	0.00
6	Toxaphene-5	34.079	33.700 E3	1.1	99	0.00
7 SA	Decachlorobiphenyl	1.673	1.638 E6	2.1	99	0.00

Signal #2

1 SA	Tetrachloro-m-xylene	7.322	7.320 E6	0.0	98	0.00
2	Toxaphene-1	32.862	35.918 E3	-9.3	95	0.00
3	Toxaphene-2	61.906	65.412 E3	-5.7	101	0.00
4	Toxaphene-3	112.503	113.200 E3	-0.6	96	0.00
5	Toxaphene-4	115.335	116.109 E3	-0.7	98	0.00
6	Toxaphene-5	90.651	89.784 E3	1.0	97	0.00
7 SA	Decachlorobiphenyl	5.234	5.195 E6	0.7	99	0.00

Evaluate Continuing Calibration Report - Not Found

Signal #2

(#) = Out of Range

SPCC's out = 0 CCC's out = 0