

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD090421\
 Data File : PD065567.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Sep 2021 21:24
 Operator : AR\AJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 ICVPD090421

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 04 01:41:19 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090421.M
 Quant Title : GC Extractables
 QLast Update : Sat Sep 04 01:23:47 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.496	4.014	50705760	376.0E6	51.740	51.306
28)	SA Decachlor...	8.239	9.085	85797195	273.3E6	49.764	51.213
Target Compounds							
2)	A alpha-BHC	3.935	4.409	77012589	549.1E6	53.456	51.907
3)	MA gamma-BHC...	4.218	4.695	74929364	510.5E6	51.911	51.805
4)	MA Heptachlor	4.517	5.199	78661451	463.8E6	51.829	51.005
5)	MB Aldrin	4.764	5.504	75428734	428.8E6	52.781	50.828
6)	B beta-BHC	4.466	4.871	38810264	238.2E6	54.439	51.170
7)	B delta-BHC	4.671	5.088	80354169	479.8E6	53.881	51.603
8)	B Heptachlo...	5.206	5.894	75449925	352.9E6	51.727	48.999
9)	A Endosulfan I	5.545	6.252	73518752	361.8E6	51.866	51.369
10)	B gamma-Chl...	5.431	6.127	78721526	386.0E6	51.541	51.397
11)	B alpha-Chl...	5.490	6.199	77873168	377.7E6	51.731	51.284
12)	B 4,4'-DDE	5.652	6.357	74817540	374.8E6	52.470	51.882
13)	MA Dieldrin	5.788	6.510	81705302	390.2E6	52.799	51.714
14)	MA Endrin	6.045	6.729	72779611	318.7E6	52.931	53.419
15)	B Endosulfa...	6.317	6.942	73844470	333.0E6	51.415	51.942
16)	A 4,4'-DDD	6.167	6.852	67485061	322.2E6	52.935	51.609
17)	MA 4,4'-DDT	6.407	7.149	72265612	318.0E6	52.504	52.212
18)	B Endrin al...	6.486	7.067	65002956	286.8E6	51.535	52.496
19)	B Endosulfa...	6.698	7.289	79287788	332.2E6	51.314	51.960
20)	A Methoxychlor	6.954	7.609	44224704	171.2E6	52.741	52.581
21)	B Endrin ke...	7.192	7.763	96397186	389.8E6	51.844	51.529
22)	Mirex	7.377	8.211	77989610	270.9E6	49.748	51.221

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD090421\
Data File : PD065567.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Sep 2021 21:24
Operator : AR\AJ
Sample : PSTDICV050
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
ICVPD090421

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 04 01:41:19 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090421.M
Quant Title : GC Extractables
QLast Update : Sat Sep 04 01:23:47 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

