

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD090424\
 Data File : PD084901.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Sep 2024 15:19
 Operator : AR\AJ
 Sample : INDA308
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3172

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/05/2024
 Supervised By : Ankita Jodhani 09/05/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 05 02:17:14 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 27 02:47:48 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.538	2.905	20481080	74016022	20.250	20.916
27) SA Decachlor...	9.032	8.105	50143148	144.1E6	38.394	41.455
Target Compounds						
2) A alpha-BHC	3.983	3.416	33608387	107.1E6	19.768	21.117
3) MA gamma-BHC...	4.312	3.751	32479580	102.6E6	20.076	21.168
4) MA Heptachlor	4.910	4.105	32406155	93285099	21.062	21.913
9) A Endosulfan I	6.052	5.271	27556639	84131351	20.030	21.948m
13) MA Dieldrin	6.325	5.539	57701442	182.1E6	39.117	42.382
14) MA Endrin	6.551	5.816	48174228	163.7E6	42.694	44.205
16) A 4,4'-DDD	6.682	5.957	39693130	149.6E6	40.202	44.091
17) MA 4,4'-DDT	6.997	6.212	35500134	133.6E6	39.659	41.754
20) A Methoxychlor	7.468	6.784	106.1E6	346.2E6	203.736	209.301

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD080424\
Data File : PD084901.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Sep 2024 15:19
Operator : AR\AJ
Sample : INDA308
Misc :
ALS Vial : 11 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 05 02:17:14 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 27 02:47:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

