

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD090425\
 Data File : PD090160.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Sep 2025 19:30
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 05 02:05:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081825.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 19 13:16:54 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.549	2.878	151.8E6	1166.8E6	51.736	52.916
28)	SA Decachlor...	9.070	8.065	226.6E6	1278.1E6	52.506	49.373
Target Compounds							
2)	A alpha-BHC	3.998	3.390	333.8E6	1897.2E6	55.606	55.901
3)	MA gamma-BHC...	4.329	3.726	314.7E6	1743.5E6	54.959	55.934
4)	MA Heptachlor	4.928	4.079	301.7E6	1690.8E6	53.794	54.207
5)	MB Aldrin	5.269	4.364	295.7E6	1690.4E6	54.614	55.504
6)	B beta-BHC	4.516	4.022	117.9E6	722.2E6	52.554	54.180
7)	B delta-BHC	4.764	4.258	306.6E6	1752.7E6	59.222	55.553
8)	B Heptachlo...	5.689	4.868	259.3E6	1508.9E6	52.847	54.484
9)	A Endosulfan I	6.073	5.241	242.6E6	1283.3E6	52.580	48.941
10)	B gamma-Chl...	5.944	5.120	262.0E6	1620.7E6	53.560	54.734
11)	B alpha-Chl...	6.025	5.185	258.4E6	1542.2E6	52.500	54.040
12)	B 4,4'-DDE	6.194	5.369	238.9E6	1560.5E6	54.436	54.142
13)	MA Dieldrin	6.346	5.507	258.3E6	1598.7E6	52.755	54.587
14)	MA Endrin	6.573	5.784	219.6E6	1491.3E6	52.986	55.997
15)	B Endosulfa...	6.785	6.075	222.2E6	1336.6E6	52.663	52.840
16)	A 4,4'-DDD	6.704	5.923	201.8E6	1320.1E6	58.333	54.903
17)	MA 4,4'-DDT	7.020	6.177	183.0E6	1253.8E6	48.044	50.014
18)	B Endrin al...	6.913	6.253	152.0E6	951.6E6	53.009	55.262
19)	B Endosulfa...	7.149	6.476	195.4E6	1252.1E6	49.859	51.357
20)	A Methoxychlor	7.492	6.748	104.6E6	711.6E6	52.159	55.182
21)	B Endrin ke...	7.629	6.985	211.4E6	1343.9E6	51.011	50.438
22)	Mirex	8.112	7.179	146.8E6	994.3E6	46.550	48.138

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD090425\
Data File : PD090160.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Sep 2025 19:30
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 05 02:05:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081825.M
Quant Title : GC Extractables
QLast Update : Tue Aug 19 13:16:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

