

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD090518\
 Data File : PD049129.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 Sep 2018 23:33
 Operator : AJ\SJ
 Sample : PIBLK21
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PIBLK21

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 06 02:11:16 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081618CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Aug 16 13:32:25 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.369	3.967	23653283	352.4E6	19.379	20.042
27)	SA Decachlor...	8.063	8.991	47952455	517.8E6	33.303	31.987
Target Compounds							
4)	MA Heptachlor	0.000	5.091f	0	2809736	N.D.	0.114
6)	B beta-BHC	0.000	4.762f	0	9942471	N.D.	0.948
8)	B Heptachlo...	0.000	5.813	0	15742472	N.D.	0.722
10)	B trans-Chl...	0.000	6.022f	0	12257074	N.D.	0.529
15)	B Endosulfa...	0.000	6.800f	0	6484655	N.D.	0.340
16)	A 4,4'-DDD	0.000	6.800	0	6484655	N.D.	0.428
20)	A Methoxychlor	0.000	7.528	0	61743851	N.D.	9.019
25)	Toxaphene-4	0.000	7.340	0	3112389	N.D.	26.524
26)	Toxaphene-5	6.913	7.528	4529570	61743851	133.428	317.371 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD090518\
Data File : PD049129.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Sep 2018 23:33
Operator : AJ\SJ
Sample : PIBLK21
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PIBLK21

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 06 02:11:16 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081618CLP.M
Quant Title : GC Extractables
QLast Update : Thu Aug 16 13:32:25 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.369 min
Delta R.T.: 0.000 min
Response: 23653283
Conc: 19.38 ng/ml

Instrument :
ECD_D
ClientSampleId :
PIBLK21




#1 Tetrachloro-m-xylene

R.T.: 3.967 min
Delta R.T.: -0.001 min
Response: 352392400
Conc: 20.04 ng/ml




#4 Heptachlor

R.T.: 0.000 min
Exp R.T.: 4.367 min
Response: 0
Conc: N.D.



#4 Heptachlor


R.T.: 5.091 min
Delta R.T.: -0.050 min
Response: 2809736
Conc: 0.11 ng/ml



#6 beta-BHC


R.T.: 0.000 min
Exp R.T.: 4.327 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
PIBLK21




#6 beta-BHC

R.T.: 4.762 min
Delta R.T.: -0.027 min
Response: 9942471
Conc: 0.95 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T.: 5.048 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 5.813 min
Delta R.T.: -0.012 min
Response: 15742472
Conc: 0.72 ng/ml



#10 trans-Chlordane


R.T.: 0.000 min
Exp R.T.: 5.270 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
PIBLK21




#10 trans-Chlordane

R.T.: 6.022 min
Delta R.T.: -0.036 min
Response: 12257074
Conc: 0.53 ng/ml




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T.: 6.149 min
Response: 0
Conc: N.D.



#15 Endosulfan II


R.T.: 6.800 min
Delta R.T.: -0.051 min
Response: 6484655
Conc: 0.34 ng/ml



#16 4,4'-DDD


R.T.: 0.000 min
Exp R.T.: 6.008 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
PIBLK21




#16 4,4'-DDD

R.T.: 6.800 min
Delta R.T.: 0.032 min
Response: 6484655
Conc: 0.43 ng/ml




#20 Methoxychlor

R.T.: 0.000 min
Exp R.T.: 6.795 min
Response: 0
Conc: N.D.



#20 Methoxychlor


R.T.: 7.528 min
Delta R.T.: 0.008 min
Response: 61743851
Conc: 9.02 ng/ml



#25 Toxaphene-4


R.T.: 0.000 min
Exp R.T.: 6.426 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :
PIBLK21




#25 Toxaphene-4

R.T.: 7.340 min
Delta R.T.: 0.058 min
Response: 3112389
Conc: 26.52 ng/ml




#26 Toxaphene-5

R.T.: 6.913 min
Delta R.T.: 0.005 min
Response: 4529570
Conc: 133.43 ng/ml



#26 Toxaphene-5


R.T.: 7.528 min
Delta R.T.: -0.065 min
Response: 61743851
Conc: 317.37 ng/ml



#27 Decachlorobiphenyl

R.T.: 8.063 min
Delta R.T.: -0.003 min
Response: 47952455
Conc: 33.30 ng/ml

Instrument :
ECD_D
ClientSampleId :
PIBLK21



#27 Decachlorobiphenyl

R.T.: 8.991 min
Delta R.T.: -0.002 min
Response: 517810715
Conc: 31.99 ng/ml