

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD090519\
 Data File : PD054418.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 Sep 2019 10:46
 Operator : SG\AJ
 Sample : K4637-06
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 06 06:56:53 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090419CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Sep 05 08:40:10 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds


Target Compounds						
15) B	Endosulfa...	0.000	6.823f	0	2973599	2.386
16) A	4,4'-DDD	0.000	6.823f	0	2973599	2.790

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD090519\
Data File : PD054418.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Sep 2019 10:46
Operator : SG\AJ
Sample : K4637-06
Misc :
ALS Vial : 4 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 06 06:56:53 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090419CLP.M
Quant Title : GC Extractables
QLast Update : Thu Sep 05 08:40:10 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm




Instrument :
ECD_D
ClientSampled :



#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T. : 3.302 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 3.973 min
Response: 0
Conc: N.D.




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T. : 6.073 min
Response: 0
Conc: N.D.



#15 Endosulfan II


R.T.: 6.823 min
Delta R.T.: -0.048 min
Response: 2973599
Conc: 2.39 ng/ml



#16 4,4'-DDD


R.T.: 0.000 min
Exp R.T.: 5.935 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#16 4,4'-DDD

R.T.: 6.823 min
Delta R.T.: 0.038 min
Response: 2973599
Conc: 2.79 ng/ml



#27 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T.: 7.993 min
Response: 0
Conc: N.D.



#27 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T.: 9.010 min
Response: 0
Conc: N.D.