

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD090523\
 Data File : PD077679.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Sep 2023 00:23
 Operator : AR\AJ
 Sample : 04250-22MSD
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 BH4C6MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 06 05:32:40 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081423CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 15 01:53:32 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.542	2.761	55679763	31554536	25.362	25.330
27)	SA Decachlor...	9.082	7.899	54103510	27183756	21.262	21.808
Target Compounds							
2)	A alpha-BHC	3.999	3.262	273.5E6	139.9E6	75.705	76.103
3)	MA gamma-BHC...	4.332	3.592	246.9E6	126.9E6	72.347	73.550
4)	MA Heptachlor	4.923	3.931	250.4E6	130.5E6	71.578	75.751
5)	MB Aldrin	5.266	4.210	249.2E6	126.4E6	71.936	73.055
6)	B beta-BHC	4.529	3.889	98919614	51363933	63.741	62.918
7)	B delta-BHC	4.777	4.118	247.5E6	128.5E6	74.660	73.963
8)	B Heptachlo...	5.693	4.712	220.0E6	118.3E6	68.943	72.341
9)	A Endosulfan I	6.080	5.082	206.7E6	108.7E6	73.138	77.165
10)	B trans-Chl...	5.950	4.962	217.9E6	115.0E6	70.558	74.488
11)	B cis-Chlor...	6.029	5.027	220.4E6	114.3E6	68.843	70.975
12)	B 4,4'-DDE	6.202	5.215	427.9E6	225.4E6	149.392	165.580
13)	MA Dieldrin	6.356	5.347	466.1E6	248.5E6	149.520	156.675
14)	MA Endrin	6.587	5.623	385.7E6	217.0E6	151.685	156.351
15)	B Endosulfa...	6.807	5.917	387.0E6	207.9E6	136.943	155.382
16)	A 4,4'-DDD	6.720	5.770	340.9E6	177.8E6	156.838	178.760
17)	MA 4,4'-DDT	7.035	6.021	372.0E6	191.9E6	168.633	185.370
18)	B Endrin al...	6.937	6.097	279.3E6	148.7E6	129.757	136.897
19)	B Endosulfa...	7.172	6.320	351.3E6	190.6E6	141.157	152.265
20)	A Methoxychlor	7.511	6.596	934.0E6	462.2E6	726.012	819.436
21)	B Endrin ke...	7.658	6.826	408.1E6	228.6E6	144.988	155.325

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD090523\
Data File : PD077679.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Sep 2023 00:23
Operator : AR\AJ
Sample : 04250-22MSD
Misc :
ALS Vial : 44 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 06 05:32:40 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081423CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 15 01:53:32 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

