

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD090723\
 Data File : PD077752.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Sep 2023 14:23
 Operator : AR\AJ
 Sample : INDA321
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3222

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/08/2023
 Supervised By : Ankita Jodhani 09/08/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 08 00:34:27 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090623CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Sep 06 16:40:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.542	2.761	47786731	30842088	18.831	20.869
27)	SA Decachlor...	9.083	7.899	99136442	60243574	31.735	37.198
Target Compounds							
2)	A alpha-BHC	3.998	3.262	78663620	46631632	18.372	20.572
3)	MA gamma-BHC...	4.330	3.591	71539445	44123624	17.694m	20.436
4)	MA Heptachlor	4.922	3.930	72980680	45945627	17.464	20.666
9)	A Endosulfan I	6.078	5.082	52553991	36364408	15.656m	20.259 #
13)	MA Dieldrin	6.355	5.347	121.9E6	82310842	32.573	39.833
14)	MA Endrin	6.586	5.623	97435007	70547705	31.581	38.586
16)	A 4,4'-DDD	6.720	5.770	90185014	58320730	33.906	43.331 #
17)	MA 4,4'-DDT	7.035	6.021	88048126	56568693	32.137	40.024
20)	A Methoxychlor	7.511	6.596	255.4E6	159.4E6	163.721	204.249

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD090723\
Data File : PD077752.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Sep 2023 14:23
Operator : AR\AJ
Sample : INDA321
Misc :
ALS Vial : 19 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 08 00:34:27 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090623CLP.M
Quant Title : GC Extractables
QLast Update : Wed Sep 06 16:40:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Instrument :
ECD_D
ClientSampled :
INDA3222

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 09/08/2023
Supervised By :Ankita Jodhani 09/08/2023