

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD090825\
 Data File : PD090184.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Sep 2025 10:26
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 09 01:40:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081825.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 19 13:16:54 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.552	2.877	151.8E6	1141.8E6	51.736	51.783
28)	SA Decachlor...	9.071	8.064	193.9E6	1099.4E6	44.923	42.470
Target Compounds							
2)	A alpha-BHC	4.002	3.389	337.3E6	1826.6E6	56.193	53.821
3)	MA gamma-BHC...	4.333	3.725	316.3E6	1683.8E6	55.236	54.019
4)	MA Heptachlor	4.930	4.078	291.8E6	1564.4E6	52.023	50.156
5)	MB Aldrin	5.272	4.363	297.8E6	1636.7E6	54.997	53.739
6)	B beta-BHC	4.519	4.021	119.8E6	708.9E6	53.432	53.179
7)	B delta-BHC	4.767	4.257	306.5E6	1684.3E6	59.205	53.385
8)	B Heptachlo...	5.691	4.867	260.9E6	1470.5E6	53.168	53.095
9)	A Endosulfan I	6.075	5.241	241.5E6	1390.7E6	52.328	53.035
10)	B gamma-Chl...	5.947	5.119	265.2E6	1589.8E6	54.216	53.690
11)	B alpha-Chl...	6.027	5.184	259.6E6	1515.7E6	52.729	53.111
12)	B 4,4'-DDE	6.197	5.369	229.2E6	1523.8E6	52.235	52.869
13)	MA Dieldrin	6.347	5.506	256.9E6	1539.4E6	52.483	52.564
14)	MA Endrin	6.575	5.782	190.6E6	1195.6E6	46.007	44.892
15)	B Endosulfa...	6.787	6.073	213.8E6	1301.7E6	50.667	51.462
16)	A 4,4'-DDD	6.705	5.923	183.6E6	1278.9E6	53.068	53.189
17)	MA 4,4'-DDT	7.021	6.176	162.4E6	1119.2E6	42.618	44.644
18)	B Endrin al...	6.915	6.252	148.5E6	909.5E6	51.784	52.818
19)	B Endosulfa...	7.150	6.475	193.6E6	1211.5E6	49.401	49.695
20)	A Methoxychlor	7.494	6.747	88500712	586.4E6	44.126	45.475
21)	B Endrin ke...	7.630	6.984	210.9E6	1376.9E6	50.894	51.678
22)	Mirex	8.114	7.177	147.1E6	964.7E6	46.625	46.702

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD090825\
Data File : PD090184.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Sep 2025 10:26
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 09 01:40:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD081825.M
Quant Title : GC Extractables
QLast Update : Tue Aug 19 13:16:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

