

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD090923\
 Data File : PD077816.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Sep 2023 22:12
 Operator : AR\AJ
 Sample : 04280-05MSD
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 JLAD2MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 09 06:29:12 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090623CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Sep 06 16:40:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.541	2.761	52533722	31177489	20.701	21.096
2)	SA Decachlor...	9.078	7.897	70304308	37202215	22.505	22.971
Target Compounds							
2)	A alpha-BHC	3.998	3.262	121.7E6	68421408	28.424	30.185
3)	MA gamma-BHC...	4.331	3.591	109.6E6	60653815	27.105	28.092
4)	MA Heptachlor	4.921	3.930	106.7E6	60303327	25.528	27.124
5)	MB Aldrin	5.264	4.210	99978524	56247027	23.948	25.095
6)	B beta-BHC	4.528	3.888	46947142	26475535	25.554	26.795
7)	B delta-BHC	4.776	4.118	110.9E6	61598991	27.607	27.635
8)	B Heptachlo...	5.691	4.711	101.8E6	57582186	26.347	27.414
9)	A Endosulfan I	6.078	5.082	96341798	54325409	28.701	30.265
10)	B trans-Chl...	5.948	4.961	98948393	56042076	26.576	28.064
11)	B cis-Chlor...	6.027	5.026	102.2E6	57220665	26.522	27.981
12)	B 4,4'-DDE	6.200	5.214	191.7E6	109.0E6	54.085	57.382
13)	MA Dieldrin	6.354	5.346	213.7E6	121.3E6	57.087	58.678
14)	MA Endrin	6.585	5.622	178.2E6	107.9E6	57.766	59.010
15)	B Endosulfa...	6.804	5.916	182.2E6	103.6E6	55.020	57.252
16)	A 4,4'-DDD	6.718	5.769	155.6E6	86499799	58.513	64.267
17)	MA 4,4'-DDT	7.033	6.020	168.9E6	89623958	61.656	63.412
18)	B Endrin al...	6.935	6.095	133.1E6	75755896	51.208	53.856
19)	B Endosulfa...	7.170	6.318	166.7E6	95732408	55.001	57.941
20)	A Methoxychlor	7.509	6.595	450.0E6	245.1E6	288.514	314.159
21)	B Endrin ke...	7.656	6.824	188.7E6	112.2E6	54.801	58.539

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD090923\
Data File : PD077816.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Sep 2023 22:12
Operator : AR\AJ
Sample : 04280-05MSD
Misc :
ALS Vial : 50 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 09 06:29:12 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090623CLP.M
Quant Title : GC Extractables
QLast Update : Wed Sep 06 16:40:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

