

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091020\
 Data File : PD059159.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Sep 2020 16:49
 Operator : DD\AJ
 Sample : L3855-11
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 11 04:29:21 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090420CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Sep 05 02:06:36 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.206	3.865	10362418	17606128	10.280	10.744
27)	SA Decachlor...	7.838	8.858	19441614	27224388	19.838	20.242
Target Compounds							
8)	B Heptachlo...	0.000	5.706	0	2753738	N.D.	1.372
9)	A Endosulfan I	0.000	6.103f	0	1804405	N.D.	1.062
10)	B trans-Chl...	0.000	5.934	0	1612305	N.D.	0.866
11)	B cis-Chlor...	0.000	6.017	0	1097128	N.D.	0.582
12)	B 4,4'-DDE	0.000	6.156	0	1291342	N.D.	0.727
19)	B Endosulfa...	0.000	7.100	0	2982318	N.D.	1.889
22)	Toxaphene-1	0.000	6.156	0	1291342	N.D.	141.353


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091020\
Data File : PD059159.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Sep 2020 16:49
Operator : DD\AJ
Sample : L3855-11
Misc :
ALS Vial : 30 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 11 04:29:21 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090420CLP.M
Quant Title : GC Extractables
QLast Update : Sat Sep 05 02:06:36 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene


R.T.: 3.206 min
Delta R.T.: 0.000 min
Response: 10362418
Conc: 10.28 ng/ml

Instrument :
ECD_D
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.865 min
Delta R.T.: 0.000 min
Response: 17606128
Conc: 10.74 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 4.845 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 5.706 min
Delta R.T.: -0.005 min
Response: 2753738
Conc: 1.37 ng/ml



#9 Endosulfan I


R.T.: 0.000 min
Exp R.T.: 5.171 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#9 Endosulfan I

R.T.: 6.103 min
Delta R.T.: 0.039 min
Response: 1804405
Conc: 1.06 ng/ml




#10 trans-Chlordane

R.T.: 0.000 min
Exp R.T.: 5.064 min
Response: 0
Conc: N.D.



#10 trans-Chlordane


R.T.: 5.934 min
Delta R.T.: -0.011 min
Response: 1612305
Conc: 0.87 ng/ml



#11 cis-Chlordane


R.T.: 0.000 min
Exp R.T.: 5.121 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#11 cis-Chlordane

R.T.: 6.017 min
Delta R.T.: 0.000 min
Response: 1097128
Conc: 0.58 ng/ml




#12 4,4'-DDE

R.T.: 0.000 min
Exp R.T.: 5.290 min
Response: 0
Conc: N.D.



#12 4,4'-DDE


R.T.: 6.156 min
Delta R.T.: -0.022 min
Response: 1291342
Conc: 0.73 ng/ml



#19 Endosulfan Sulfate


R.T.: 0.000 min
Exp R.T. : 6.312 min
Response: 0
Conc: N.D.

Instrument :
ECD_D
ClientSampleId :




#19 Endosulfan Sulfate

R.T.: 7.100 min
Delta R.T.: 0.008 min
Response: 2982318
Conc: 1.89 ng/ml




#22 Toxaphene-1

R.T.: 0.000 min
Exp R.T. : 5.377 min
Response: 0
Conc: N.D.



#22 Toxaphene-1


R.T.: 6.156 min
Delta R.T.: -0.067 min
Response: 1291342
Conc: 141.35 ng/ml



#27 Decachlorobiphenyl

R.T.: 7.838 min
Delta R.T.: -0.001 min
Response: 19441614
Conc: 19.84 ng/ml

Instrument :
ECD_D
ClientSampleId :



#27 Decachlorobiphenyl

R.T.: 8.858 min
Delta R.T.: -0.003 min
Response: 27224388
Conc: 20.24 ng/ml