



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091020\
Data File : PD059139.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Sep 2020 10:53
Operator : DD\AJ
Sample : L3855-05MSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 11 05:34:33 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090420CLP.M
Quant Title : GC Extractables
QLast Update : Sat Sep 05 02:06:36 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



QEdit

(1) Tetrachloro-m-xylene (SA)

3.206min 16.890 ng/ml

response 17025341

(1) Tetrachloro-m-xylene #2 (SA)


3.866min 19.709 ng/ml

response 32298321

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091020\
Data File : PD059139.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Sep 2020 10:53
Operator : DD\AJ
Sample : L3855-05MSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 11 05:34:33 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090420CLP.M
Quant Title : GC Extractables
QLast Update : Sat Sep 05 02:06:36 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



QEdit

(1) Tetrachloro-m-xylene (SA)

3.206min 16.890 ng/ml

response 17025341

(1) Tetrachloro-m-xylene #2 (SA)

3.864min 20.961 ng/ml m

response 34349246

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091020\
 Data File : PD059139.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Sep 2020 10:53
 Operator : DD\AJ
 Sample : L3855-05MSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 11 05:34:33 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090420CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Sep 05 02:06:36 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.206	3.864	17025341	34349246	16.890	20.961m
27) SA Decachlor...	7.837	8.858	29574899	44499396	30.177	33.087
Target Compounds						
3) MA gamma-BHC...	3.899	4.528	77285879	134.9E6	48.795	61.224 #
4) MA Heptachlor	4.177	5.027	58858450	110.8E6	52.406	52.572
5) MB Aldrin	4.412	5.327	71190928	108.1E6	54.694	52.389
13) MA Dieldrin	5.408	6.318	116.9E6	198.4E6	103.881	107.151
14) MA Endrin	5.667	6.534	308.6E6	165.7E6	332.964	102.776 #
17) MA 4,4'-DDT	6.028	6.959	94191007	163.0E6	111.088	111.135

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

AJ
 09/19/20