

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091120\
 Data File : PD059171.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Sep 2020 09:57
 Operator : DD\AJ
 Sample : MDL-S-04
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 MDL-S-04

Manual Integrations
 APPROVED

Ankita
 9/14/2020 11:05:10 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 11 10:29:27 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090420CLP.M
 Quant Title : GC Extractables
 QLast Update : Sat Sep 05 02:06:36 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.205	3.864	19461028	34821429	19.306	21.249
27)	SA Decachlor...	7.838	8.857	37334126	53589888	38.095	39.846
Target Compounds							
2)	A alpha-BHC	3.627	4.247	1355267	2440171	0.864	1.005
3)	MA gamma-BHC...	3.898	4.527	1519610	2044357	0.959	0.928
4)	MA Heptachlor	4.176	5.027	1161988	2066626	1.035	0.981
5)	MB Aldrin	4.411	5.326	1303984	1965715	1.002	0.953
6)	B beta-BHC	4.147	4.700	665134	948437	1.084	0.996
7)	B delta-BHC	4.337	4.909	1293416	2018512	0.889	1.005
8)	B Heptachlo...	4.843	5.710	1099834	2301062	0.963	1.146
9)	A Endosulfan I	5.170	6.063	1114386	2099295	1.044	1.235
10)	B trans-Chl...	5.062	5.943	1208461	1932753	1.012	1.038
11)	B cis-Chlor...	5.119	6.015	1184746	2152815	0.993	1.143
12)	B 4,4'-DDE	5.288	6.175	1065886	2150804	0.899	1.211 #
13)	MA Dieldrin	5.407	6.318	1119532	1922057	0.995	1.038
14)	MA Endrin	5.658	6.534	854471	1298511	0.922	0.806
15)	B Endosulfa...	5.932	6.744	976219	1683274	0.952	1.089
16)	A 4,4'-DDD	5.795	6.662	867372	1304386	1.090	0.914
17)	MA 4,4'-DDT	6.028	6.959	656731	1041024	0.775	0.710
18)	B Endrin al...	6.098	6.868	571304	1468670	0.725	1.094 #
19)	B Endosulfa...	6.309	7.090	918315	1552293	0.950	0.983
20)	A Methoxychlor	6.575	7.418	241891	674825	0.679m	0.875m#
21)	B Endrin ke...	6.790	7.555	1164731	1708213	1.085	1.002

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091120\
Data File : PD059171.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Sep 2020 09:57
Operator : DD\AJ
Sample : MDL-S-04
Misc :
ALS Vial : 8 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 11 10:29:27 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090420CLP.M
Quant Title : GC Extractables
QLast Update : Sat Sep 05 02:06:36 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Manual Integrations
APPROVED

Ankita
9/14/2020 11:05:10 AM

Instrument :
ECD_D
Client Sample :
MDL-S-04