

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091123\
 Data File : PD077833.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Sep 2023 11:51
 Operator : AR\AJ
 Sample : INDA301
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA3061

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 11 12:20:22 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091123CLP-TCLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Sep 11 12:16:08 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.541	2.759	50667133	31452197	20.000	20.000
27)	SA Decachlor...	9.084	7.899	126.0E6	68731743	40.000	40.000
Target Compounds							
2)	A alpha-BHC	3.998	3.261	84632485	48722027	20.000	20.000
3)	MA gamma-BHC...	4.331	3.590	79248870	45881047	20.000	20.000
4)	MA Heptachlor	4.922	3.929	82686124	46920626	20.000	20.000
9)	A Endosulfan I	6.080	5.082	66760708	38265818	20.000	20.000
13)	MA Dieldrin	6.355	5.347	149.0E6	89224612	40.000	40.000
14)	MA Endrin	6.586	5.623	118.5E6	76271915	40.000	40.000
16)	A 4,4'-DDD	6.720	5.769	104.5E6	57440118	40.000	40.000
17)	MA 4,4'-DDT	7.035	6.021	105.1E6	59606295	40.000	40.000
20)	A Methoxychlor	7.511	6.596	301.5E6	164.1E6	200.000	200.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091123\
Data File : PD077833.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Sep 2023 11:51
Operator : AR\AJ
Sample : INDA301
Misc :
ALS Vial : 14 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 11 12:20:22 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091123CLP-TCLP.M
Quant Title : GC Extractables
QLast Update : Mon Sep 11 12:16:08 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

