

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091219\
 Data File : PD054548.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Sep 2019 5:36
 Operator : SG\AJ
 Sample : INDA347
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 INDA347

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 13 06:30:23 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090419CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu Sep 05 08:40:10 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.301	3.972	15593760	21851845	22.063	22.141
27)	SA Decachlor...	7.988	9.008	29980022	37713611	44.300	40.314
Target Compounds							
2)	A alpha-BHC	3.730	4.359	22434158	30498621	20.646	20.451
3)	MA gamma-BHC...	4.006	4.642	20988999	29135781	21.272	20.794
4)	MA Heptachlor	4.293	5.149	19253219	29571560	25.739	21.451
9)	A Endosulfan I	5.303	6.191	17500718	24845143	20.510	19.988
13)	MA Dieldrin	5.544	6.446	39466640	53996615	42.248	40.056
14)	MA Endrin	5.797	6.662	26977786	32940790	35.128	32.241
16)	A 4,4'-DDD	5.931	6.783	30660576	43274182	44.295	40.607
17)	MA 4,4'-DDT	6.168	7.081	26692613	39978573	42.424	38.477
20)	A Methoxychlor	6.717	7.538	72166890	110.8E6	257.423	211.264

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091219\
Data File : PD054548.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Sep 2019 5:36
Operator : SG\AJ
Sample : INDA347
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
INDA347

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 13 06:30:23 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD090419CLP.M
Quant Title : GC Extractables
QLast Update : Thu Sep 05 08:40:10 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

