

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091324\
 Data File : PD085096.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Sep 2024 18:59
 Operator : AR\AJ
 Sample : P3898-07
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 DDC57

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/18/2024
 Supervised By : Ankita Jodhani 09/18/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 14 00:53:21 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 27 02:47:48 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds						
1) SA Tetrachlo...	3.527	2.915	37755548	2635.0E6	37.330m	744.597m#
2) SA Decachlor...	9.035	8.104	27091793	69268665	20.744	19.921
Target Compounds						
2) A alpha-BHC	3.981	3.412	3507131	14403668	2.063m	2.841m#
9) A Endosulfan I	6.056	5.281	21107384	32576111	15.342	8.498 #
10) B trans-Chl...	5.902	5.145	9384617	30198291	6.654m	7.342m
11) B cis-Chlor...	5.997	5.228	3101038	23464573	2.097m	5.593m#
14) MA Endrin	6.542	5.811	3748962	26691951	3.323m	7.209 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091324\
Data File : PD085096.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Sep 2024 18:59
Operator : AR\AJ
Sample : P3898-07
Misc :
ALS Vial : 36 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 14 00:53:21 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 27 02:47:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



Reviewed By: Abdul Miza 09/18/2024
Supervised By: Ankita Jodhani 09/18/2024

Manual Integrations
APPROVED

Instrument :
ECD_D
ClientSampled :
DDC57