

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091324\
Data File : PD085087.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Sep 2024 16:59
Operator : AR\AJ
Sample : P3877-16RXRE
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
DCVZ5RXRE

Manual IntegrationsAPPROVED

Reviewed By :Abdul

Mirza

09/15/2024


Supervised By :Ankita

Jodhani

09/16/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 14 00:50:20 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 27 02:47:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

3.536min -136.929 ng/ml

response -138490350

(1) Tetrachloro-m-xylene #2 (SA)

2.916min 92.503 ng/ml

response 327348961

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091324\
Data File : PD085087.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Sep 2024 16:59
Operator : AR\AJ
Sample : P3877-16RXRE
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
DCVZ5RXRE

Manual IntegrationsAPPROVED

Reviewed By :Abdul

Mirza

09/15/2024


Supervised By :Ankita

Jodhani

09/16/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 14 00:50:20 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 27 02:47:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

3.535min 21.272 ng/ml m

response 21514711

(1) Tetrachloro-m-xylene #2 (SA)

2.915min 78.451 ng/ml m

response 277621604

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091324\
 Data File : PD085087.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Sep 2024 16:59
 Operator : AR\AJ
 Sample : P3877-16RXRE
 Misc :
 ALS Vial : 27 Sample Multiplier: 1


Instrument :
 ECD_D
 ClientSampleId :
 DCVZ5RXRE

Manual IntegrationsAPPROVED

Reviewed By :Abdul
 Mirza
 09/15/2024
 Supervised By :Ankita
 Jodhani
 09/16/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 14 00:50:20 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 27 02:47:48 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(9) Endosulfan I (A)

6.067min 13.414 ng/ml

response 18454857

(9) Endosulfan I #2 (A)

5.278min 2.804 ng/ml

response 10748074

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091324\
Data File : PD085087.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Sep 2024 16:59
Operator : AR\AJ
Sample : P3877-16RXRE
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
DCVZ5RXRE


Manual IntegrationsAPPROVED

Reviewed By :Abdul
Mirza
09/15/2024

Supervised By :Ankita
Jodhani
09/16/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 14 00:50:20 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 27 02:47:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(9) Endosulfan I (A)

6.061min 3.841 ng/ml m

response 5284266

(9) Endosulfan I #2 (A)

5.278min 2.804 ng/ml

response 10748074

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091324\
Data File : PD085087.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Sep 2024 16:59
Operator : AR\AJ
Sample : P3877-16RXRE
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
DCVZ5RXRE

Manual IntegrationsAPPROVED

Reviewed By :Abdul

Mirza

09/15/2024


Supervised By :Ankita

Jodhani

09/16/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 14 00:50:20 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 27 02:47:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(10) trans-Chlordane (B)

5.912min 1.075 ng/ml

response 1516159

(10) trans-Chlordane #2 (B)

5.145min 3.277 ng/ml

response 13477060

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091324\
Data File : PD085087.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Sep 2024 16:59
Operator : AR\AJ
Sample : P3877-16RXRE
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
DCVZ5RXRE

Manual IntegrationsAPPROVED

Reviewed By :Abdul

Mirza

09/15/2024



Supervised By :Ankita

Jodhani

09/16/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 14 00:50:20 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 27 02:47:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(10) trans-Chlordane (B)

5.910min 1.260 ng/ml m

response 1777182

(10) trans-Chlordane #2 (B)

5.144min 2.675 ng/ml m

response 11003759

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091324\
Data File : PD085087.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Sep 2024 16:59
Operator : AR\AJ
Sample : P3877-16RXRE
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
DCVZ5RXRE

Manual IntegrationsAPPROVED

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 14 00:50:20 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 27 02:47:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Reviewed By :Abdul

Mirza


09/15/2024

Supervised By :Ankita

Jodhani

09/16/2024

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(13) Dieldrin (MA)

6.327min 1.112 ng/ml

response 1640938

(13) Dieldrin #2 (MA)

0.000min 0.000 ng/ml

response 0

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091324\
Data File : PD085087.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Sep 2024 16:59
Operator : AR\AJ
Sample : P3877-16RXRE
Misc :
ALS Vial : 27 Sample Multiplier: 1


Instrument :
ECD_D
ClientSampleId :
DCVZ5RXRE

Manual IntegrationsAPPROVED

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 14 00:50:20 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD082524CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 27 02:47:48 2024
Response via : Initial Calibration
Integrator: ChemStation

Reviewed By :Abdul
Mirza
09/15/2024
Supervised By :Ankita
Jodhani
09/16/2024

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm



(13) Dieldrin (MA)

6.326min 1.056 ng/ml m
response 1557713

(13) Dieldrin #2 (MA)

5.544min 1.574 ng/ml m
response 6760900