

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091522\
 Data File : PD071883.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Sep 2022 10:04
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/15/2022
 Supervised By :Ankita Jodhani 09/15/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 14 23:44:25 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091022.M
 Quant Title : GC Extractables
 QLast Update : Sat Sep 10 22:11:50 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.805	3.564	137.4E6	1702.7E6	56.200	59.338
28)	SA Decachlor...	7.996	9.090	117.2E6	510.3E6	46.533	54.891
Target Compounds							
2)	A alpha-BHC	3.314	4.030	221.7E6	2372.8E6	57.533	59.129
3)	MA gamma-BHC...	3.647	4.365	202.6E6	1926.4E6	57.688	58.675
4)	MA Heptachlor	3.992	4.943	207.8E6	1389.5E6	55.578m	57.506
5)	MB Aldrin	4.276	5.285	206.6E6	1164.8E6	56.424m	57.460
6)	B beta-BHC	3.945	4.572	85274231	760.2E6	56.283	57.521
7)	B delta-BHC	4.179	4.820	210.3E6	1490.1E6	59.662	63.749
8)	B Heptachlo...	4.784	5.717	186.4E6	997.4E6	57.001	64.680
9)	A Endosulfan I	5.157	6.103	162.7E6	837.0E6	51.182	56.773m
10)	B gamma-Chl...	5.037	5.970	191.0E6	923.0E6	56.673	57.442m
11)	B alpha-Chl...	5.101	6.047	186.2E6	884.3E6	56.118	56.676m
12)	B 4,4'-DDE	5.291	6.222	173.3E6	861.4E6	56.642	58.476
13)	MA Dieldrin	5.424	6.382	182.8E6	899.7E6	54.642	59.362
14)	MA Endrin	5.701	6.614	151.4E6	639.9E6	51.112	50.664
15)	B Endosulfa...	5.996	6.840	138.9E6	686.1E6	50.441	55.411
16)	A 4,4'-DDD	5.849	6.746	141.2E6	703.1E6	55.685	59.158
17)	MA 4,4'-DDT	6.103	7.057	138.3E6	675.7E6	52.563	54.696
18)	B Endrin al...	6.177	6.971	113.7E6	547.6E6	51.394	56.716
19)	B Endosulfa...	6.402	7.204	142.2E6	639.3E6	53.024	57.321
20)	A Methoxychlor	6.681	7.534	75512795	331.0E6	50.431	54.448
21)	B Endrin ke...	6.911	7.696	163.0E6	710.2E6	54.315	57.102
22)	Mirex	7.097	8.157	118.1E6	463.5E6	50.122	53.523

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091522\
 Data File : PD071883.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Sep 2022 10:04
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 09/15/2022
 Supervised By :Ankita Jodhani 09/15/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 14 23:44:25 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091022.M
 Quant Title : GC Extractables
 QLast Update : Sat Sep 10 22:11:50 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

