

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091522\
 Data File : PD071894.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Sep 2022 12:43
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/15/2022
 Supervised By : Ankita Jodhani 09/15/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 15 00:36:41 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091022.M
 Quant Title : GC Extractables
 QLast Update : Sat Sep 10 22:11:50 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.806	3.564	138.1E6	1617.8E6	56.485	56.379
28)	SA Decachlor...	7.996	9.089	103.5E6	456.8E6	41.100	49.133
Target Compounds							
2)	A alpha-BHC	3.314	4.030	219.6E6	2205.7E6	56.974	54.966
3)	MA gamma-BHC...	3.647	4.366	198.4E6	1789.5E6	56.510	54.504
4)	MA Heptachlor	3.994	4.943	211.0E6	1280.0E6	56.448	52.976
5)	MB Aldrin	4.276	5.285	202.1E6	1074.4E6	55.190m	52.999
6)	B beta-BHC	3.946	4.572	84515248	702.3E6	55.782	53.140
7)	B delta-BHC	4.178	4.820	201.2E6	1367.5E6	57.106m	58.504
8)	B Heptachlo...	4.783	5.715	173.7E6	855.9E6	53.110	55.506m
9)	A Endosulfan I	5.157	6.103	151.6E6	758.4E6	47.698	51.447m
10)	B gamma-Chl...	5.036	5.970	175.5E6	821.5E6	52.050	51.129m
11)	B alpha-Chl...	5.101	6.049	170.8E6	845.5E6	51.453	54.188
12)	B 4,4'-DDE	5.292	6.221	160.5E6	771.7E6	52.454	52.387
13)	MA Dieldrin	5.425	6.381	169.7E6	795.4E6	50.748	52.477
14)	MA Endrin	5.702	6.613	139.8E6	559.5E6	47.179	44.296
15)	B Endosulfa...	5.997	6.840	128.8E6	610.5E6	46.757	49.302
16)	A 4,4'-DDD	5.850	6.746	131.0E6	629.7E6	51.654	52.980
17)	MA 4,4'-DDT	6.103	7.057	123.9E6	582.5E6	47.072	47.151
18)	B Endrin al...	6.177	6.971	101.3E6	489.0E6	45.800	50.651
19)	B Endosulfa...	6.402	7.203	124.8E6	561.5E6	46.534	50.345
20)	A Methoxychlor	6.682	7.533	64182794	291.5E6	42.864	47.957
21)	B Endrin ke...	6.911	7.696	143.3E6	620.5E6	47.762	49.891
22)	Mirex	7.097	8.157	103.5E6	395.0E6	43.897	45.615

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091522\
Data File : PD071894.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Sep 2022 12:43
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 09/15/2022
Supervised By :Ankita Jodhani 09/15/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 15 00:36:41 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091022.M
Quant Title : GC Extractables
QLast Update : Sat Sep 10 22:11:50 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

