

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091522\
 Data File : PD071899.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Sep 2022 16:03
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 09/15/2022
 Supervised By : Ankita Jodhani 09/15/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 15 00:39:31 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091022.M
 Quant Title : GC Extractables
 QLast Update : Sat Sep 10 22:11:50 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	2.811	3.565	130.0E6	1583.8E6	53.157	55.193
28)	SA Decachlor...	7.999	9.092	108.4E6	450.0E6	43.022	48.400
Target Compounds							
2)	A alpha-BHC	3.320	4.031	211.5E6	2212.6E6	54.874	55.136
3)	MA gamma-BHC...	3.654	4.367	192.9E6	1794.7E6	54.938	54.662
4)	MA Heptachlor	3.999	4.944	197.9E6	1277.5E6	52.931	52.872
5)	MB Aldrin	4.282	5.287	191.5E6	1094.8E6	52.304m	54.003
6)	B beta-BHC	3.952	4.573	80713398	709.8E6	53.273	53.703
7)	B delta-BHC	4.185	4.822	196.9E6	1392.7E6	55.870	59.580
8)	B Heptachlo...	4.789	5.717	170.2E6	891.0E6	52.059	57.786m
9)	A Endosulfan I	5.163	6.105	152.2E6	770.8E6	47.880	52.288m
10)	B gamma-Chl...	5.042	5.972	176.1E6	862.7E6	52.243	53.691m
11)	B alpha-Chl...	5.107	6.051	171.6E6	866.0E6	51.708	55.506
12)	B 4,4'-DDE	5.297	6.223	159.8E6	776.2E6	52.217	52.693
13)	MA Dieldrin	5.430	6.383	172.8E6	816.8E6	51.658	53.891
14)	MA Endrin	5.707	6.615	143.8E6	595.6E6	48.531	47.153
15)	B Endosulfa...	6.002	6.842	130.3E6	618.7E6	47.320	49.969
16)	A 4,4'-DDD	5.855	6.748	137.9E6	656.5E6	54.371	55.233
17)	MA 4,4'-DDT	6.108	7.059	116.2E6	558.3E6	44.133	45.193
18)	B Endrin al...	6.182	6.973	102.5E6	487.8E6	46.337	50.518
19)	B Endosulfa...	6.406	7.206	122.9E6	571.7E6	45.807	51.263
20)	A Methoxychlor	6.687	7.535	61806927	280.5E6	41.278	46.139
21)	B Endrin ke...	6.916	7.698	152.0E6	630.6E6	50.657	50.700
22)	Mirex	7.102	8.160	114.8E6	405.3E6	48.726	46.795

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD091522\
Data File : PD071899.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Sep 2022 16:03
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_D

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 09/15/2022
Supervised By :Ankita Jodhani 09/15/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 15 00:39:31 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD091022.M
Quant Title : GC Extractables
QLast Update : Sat Sep 10 22:11:50 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

